



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004, INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - JUNE 2024

LAB NO: 20400

LAB NAME: GHSPL SAMBHAJV KNJ HEALTH CARE LLP

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 22/06/2024

Date of Result Published: 05/07/2024

PT Item: Lyophilized human serum based

S. No	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1663	158.64	11.10	17.73	170.84 mg/dL	0.73	0.61
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1241	55.67	13.56	7.69	58 mg/dL	0.57	0.40
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	690	2.66	13.5	0.39	2.64 mg/dL	0.44	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1770	3.9	18.38	0.72	3.87 mg/dL	0.7	0.04
5	I-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1497	5.52	10.7	0.59	5.41 g/dL	-0.04	0.05
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1457	3.3	10.15	0.34	3.2 g/dL	0.29	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1511	7.05	14.78	1.04	5.45 mg/dL	0.58	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1854	119.63	12.35	14.77	114 mg/dL	-0.61	0.57
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1447	115.23	16.87	19.44	115 mg/dL	-0.07	0.66



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004, INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - JUNE 2024

LAB NO: 20400

LAB NAME: GHSPL SAMBAJY KNJ HEALTH CARE LLP

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 22/06/2024

Date of Result Published: 05/07/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			

u* - method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.

LAB ADDRESS

GHSPL SAMBAJY KNJ HEALTH CARE LLP
BHATJANGLA, KRISHNANIGAR, NH- 39, OPP
HOTEL HAVELI,
West Bengal

Coordinator Contact Details:

E-Mail: olmas@cmcvellore.ac.in
Contact No: 0418 2293102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any examinations.
**** End of Report ****