



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004, INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



MONTHLY SUMMARY REPORT - JULY 2024

LAB NO: 20400

LAB NAME: GHSPL SAMBHAJV KNJ HEALTH CARE LLP

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 19/07/2024

Date of Result Published: 08/08/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1863	159.64	11.11	17.73	172.73 mg/dL	0.74	0.82
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1341	56.68	13.58	7.69	61 mg/dL	0.56	0.42
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	870	2.86	13.5	0.39	2.99 mg/dL	0.33	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1670	3.9	18.38	0.72	4.33 mg/dL	0.6	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1500	5.52	10.7	0.59	5.51 g/dL	-0.02	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1459	3.3	10.15	0.34	3.4 g/dL	0.29	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1671	7.05	14.79	1.04	6.45 mg/dL	-0.55	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1829	119.63	12.35	14.77	112 mg/dL	-0.52	0.69
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1634	115.23	16.87	19.44	114 mg/dL	-0.06	0.96



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004, INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - JULY 2024

LAB NO: 20400

LAB NAME: GHSPL SAMBAJIV KNJ HEALTH CARE LLP

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 19/07/2024

Date of Result Published: 08/08/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			

u* - method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

LAB ADDRESS

GHSPL SAMBAJIV KNJ HEALTH CARE LLP,
BHATJANGLA, KRISHNANAGAR, NH- 39, OPP
HOTEL HAVELI,
West Bengal

Coordinator Contact Details:

E-Mail: clinqc@cmcvellore.ac.in
Contact No.: 0476-2283162

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components.
***** End of Report *****

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.