



# CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004, INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

## MONTHLY SUMMARY REPORT - AUGUST 2024

LAB NO: 20400

LAB NAME: GHSPL SAMBAJY KNJ HEALTH CARE LLP

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 20/08/2024

Date of Result Published: 02/09/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1884	73.06	12.16	8.88	69.7 mg/dL	-0.38	0.41
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1355	90.77	16.23	14.74	101.4 mg/dL	0.72	0.8
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	689	1.29	13.11	0.17	1.31 mg/dL	0.12	0.01
4	T.BILIRUBIN II	Diazonium Salt (Cokrimetric) / Jendressik II	Any Analyser (Automation / Semi Automation)	1700	4.18	20.3	0.85	4.94 mg/dL	0.88	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1530	6.41	11.3	0.61	5.34 g/dL	-0.11	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1487	3.26	10.57	0.34	3.37 g/dL	0.02	0.02
7	URIC ACID II	Enzymatic / Uricase Cokrimetric II	Any Analyser (Automation / Semi Automation)	1692	5.17	20.67	1.07	4.1 mg/dL	-1	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1852	111.8	12.69	14.19	103.72 mg/dL	-0.57	0.66
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1653	200.38	14.08	28.21	215.88 mg/dL	0.66	1.39



# CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004, INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME



## MONTHLY SUMMARY REPORT - AUGUST 2024

LAB NO: 20407

LAB NAME: **GHSPL SAMBAJY KNJ HEALTH CARE LLP**

CONSTITUENT GROUP

**Chemistry II**

Date of Result Entered: 20/08/2024

Date of Result Published: 02/09/2024

PT Item: Lyophilized human serum bases

S. No.	Analyte	Method / Principle	Analyser	No. of Tests	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			

u\* - method of uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  > 3.0$	Unacceptable (Action Signal)

### LAB ADDRESS

GHSPL SAMBAJY KNJ HEALTH CARE LLP,  
BHALUANGI A, KRISHNANIGAR, NH-39, OPP  
HOTEL IVEELI,  
West Bengal

### Coordinator Contact Details:

E-Mail: [cmq@cmcvellore.edu.in](mailto:cmq@cmcvellore.edu.in)  
Contact No.: 0416-2293102

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is as follows.  
CMC EQAS does not hold contract any comments.  
\*\*\* End of Report \*\*\*

Dr. Pamela Christudoss,  
CMC EQAS Coordinator,  
Christian Medical College, Vellore.