



FINESTAR CALIBRATION PVT. LTD.

Plot No. 280, Sector No. 21, Yamunanagar, PCNTDA, Nigdi, Pune - 411 044 (MH)
Mob. : 7276470703, Email : finecalibration@gmail.com

CALIBRATION CERTIFICATE

1) Customer

Name & Address : ADVANCE-TECH PATHOLOGY LAB
KAZI COMPLEX, OLD PUNE-NASHIK HIGHWAY,
MAHATMA PHULE CHOWK, CHAKAN - 410501

Calibration Certificate No. : FSC/2408/0501/004

Date of Received : 05.08.2024
Date of Calibration : 05.08.2024
Calibration Due On : 04.08.2025
Date of Issue : 06.08.2024
Location of Calibration : On Site
Work Instruction No : MICROSWI/01

Amb. Temp. : 25.9 °C
Relative Humidity : 55 %RH
Condition of Item : Physically Ok

2) Details of UUC

Name : Microscope
ID No : MS-01
Make : Micro Lux

Range : 10,40,100 x

3) Standard Used for Calibration

Name : Linear Glass Scale
Certificate No : CH/24/0308P-04
Certified By : Calibration House
ID/Sr No : FSC-LGS-01
Due Date : 02.08.2025


4) Calibration Results :

Calibration Point	Standard Readings	UUC Readings	Error	Expanded Uncertainty
X	X	X	X	±%
10	10	10	0	1.50
40	40	40	0	1.50
100	100	100	0	1.50

Note :


- The reported Expanded Uncertainty is calculated at 95.45 % Confidence level with coverage factor k=2.
- Equipment used for calibration were calibrated & traceable to National & International Standards.
- This report refers only to particular item (s) submitted for calibration. UUC stands for Unit Under Calibration.
- The calibration results reported are valid at the time of and under the stated conditions of the measurements.
- This report shall not be reproduced in full / part without prior permission of Finestar Calibration Pvt Ltd.
- The value measured of uuc & standard are mean of 5 reading.

Calibrated By


Calibration Engineer
Abhijeet S.



Approved By


Quality Manager
Mr Shaik Chorgaste

RF-550

** End of the certificate **

Page No 1 of 1