

CALIBRATED ACCURATE MEASUREMENTS
CERT No. AM24CC04502
ID No. 360050354
CAL Dt. 01.08.2024 SIGN RA
SUGG DUE Dt. 31.07.2025
9811826626

CALIBRATION CERTIFICATE



AM-F-12

PAGE 1 OF 1

ASAGAR STATE GENERAL HOSPITAL

4, BRAMHO SAMAJ ROAD,
KOLKATA - 700034

ADDRESS

ITEM DESCRIPTION : **CENTRIFUGE**
Sr. No. / ID No. : 360050354/360050354
MAKE / MODEL : REMI/---
RANGE / SIZE : 5000 RPM
LEAST COUNT : REFER OBSERVATION
LOCATION : ---

CERTIFICATE No. : AM24CC04502
ULR No. : CC228624000004502F
ENVIRONMENTAL : TEMPERATURE (20 ± 2)°C
CONDITIONS : HUMIDITY (60 ± 20) % RH
CALIB. PROCEDURE : AM-WI-57M
CALIB. PERFORMED AT : LAB
RECEIPT DATE : 31/07/2024
CALIBRATION DATE : 01/08/2024
SUGG. DUE DATE : 31/07/2025
ISSUE DATE : 03/08/2024
CONDITION ON RECEIPT : SATISFACTORY

TRACEABILITY :-

STANDARDS USED FOR CALIBRATION ARE TRACEABLE TO NATIONAL OR INTERNATIONAL STANDARDS THROUGH NABL ACCREDITED LABORATORY.

MASTER(S) USED FOR CALIBRATION :-

Sr. No.	ITEM No.	DESCRIPTION	CALIB. DUE DATE	CERTIFICATE No.	CALIBRATED FROM
1	H.364908 / AM/TM 001	DIGITAL TACHOMETER	28/06/2025	AM24CC03721	ACCURATE MEASUREMENTS, DELHI

I. MECHANICAL CALIBRATION

OBSERVATIONS

Sr. No.	STANDARD (in RPM)	UNIT UNDER CALIBRATION (in RPM)	ERROR (in RPM)	Uncertainty at 95.45% C.L, k=2 (in RPM)
1	4997	5000	3	4.2

NON CONTACT



REMARK :-

END OF CERTIFICATE

Calibrated By: - (Vikas Sharma, Calib. Engg.)

Approved By: (S. Kaushik, QM)

- NOTE :-
- 1 The calibration result reported in this certificate are related to item (s) calibrated & is valid at the time of and under stated condition of measurement.
 - 2 This certificate should not be reproduced, expect in full, without our prior permission in writing.
 - 3 Calibration done by the laboratory is meant for scientific and industrial purpose only.

ACCURATE MEASUREMENTS, PLOT NO. 103, D BLOCK, MEETHAPUR EXTENSION, BADARPUR, NEW DELHI - 110044

E-Mail : accurate2@gmail.com ; Mob. No. 9818166269, 9811826626

ISSUE/REVISION NO. : 04/04

REVISION DATE : 01/05/23