

Urine Routine Analysis

METHOD WISE REPORT PART-01

Cycle - 13/2024

Date: 28/09/2024

Lab Code: 3345

Round No -09

Parameters	Units	Method Group	No. of Participants	Consensus		Consensus Result	Your Value	Remark
				Number	%			
Specific Gravity	-	Semi-Automated	205	133	65%	1.020	1.020	Within Consensus
pH	-	Semi-Automated	205	135	66%	7.0	6.5	Within Consensus
Leukocytes	/µl	Semi-Automated	205	168	82%	2+-3+	2+	Within Consensus
Nitrite	-	Semi-Automated	204	181	89%	Positive	Positive	Within Consensus
Protein	mg/dl	Semi-Automated	205	166	81%	2+-3+	2+	Within Consensus
Glucose	mg/dl	Semi-Automated	205	147	72%	3+-4+	3+	Within Consensus
Ketones	-	Semi-Automated	205	138	67%	1+-2+	1+	Within Consensus
Urobilinogen	mg/dl	Semi-Automated	205	148	72%	2+-3+	3+	Within Consensus
Bilirubin	mg/dl	Semi-Automated	205	156	76%	3+	3+	Within Consensus
Blood (Hemoglobin)	/µl	Semi-Automated	205	161	79%	3+-4+	3+	Within Consensus

Legends	≥ 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#)Late Result Submission
---------	-----------------	-----------------------	------------------	------------------	---------------------------

Chief Coordinator



Dr. Sanjay Mehrotra

Prepared By: SG

Programme Director



Dr. Bandana Mehrotra

Checked By: 