



WE ALWAYS TRY TO GET BEST MEASUREMENT

सूर्यआदर्श कैलिब्रेशन एंड कंसल्टेंट (ओपीसी) प्राइवेट लिमिटेड

Formerly UNIQUE CALIBRATION & CONSULTANT SERVICES
AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY
NABL Accredited Calibration Laboratory As per ISO/IEC 17025:2017



CALIBRATION CERTIFICATE

Format No.: SCCPL-QF-7.8.1		Certificate No.: SCCPLCC20240916-003		ULR No.: CC328724000013111F	
Group: Volume		Discipline: Mechanical			
Customer's Name & Address			Page No.: 1 of 1		
M/s. Sood Diagnostics			Calibration Performed: At Lab		
Shop No: 1, 165, Sadar Bazar Rd, Basantra Line, Sadar Bazaar,			Date of Receipt: 16-09-2024		
Delhi Cantonment, New Delhi, Delhi 110010			Service Request No.: SCCPL5RN20240916-003		
			Certificate Issue Date: 18-09-2024		
			Date of Calibration: 17-09-2024		
			Next Due Date (Suggested by Customer): 15-09-2025		
Equipment Description (UUC) & Identification					
Name of the Instrument: Micro Pipette			Make: Agappe-Nispa-L		
Range: 20 µl to 200 µl			Model: -----		
Least Count: 0.1 µl			Sr. No.: 0W00572		
Condition of Instrument: Ok			I.D. No.: -----		
Master Equipment Used			Location: -----		
Sr. No.	Instrument's Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	Digital Weighing Balance	0 to 220 g & SCCPL/CAL/005	SCCPL	SCCPLCC20240113-002	14-01-2025
Environmental Conditions			Calibration Procedure		Calibration Reference Standard
Temperature: 27 ± 3 °C, Humidity: 45 to 80 %RH			SCCPL 03 SOP 018		ISO 8655
Atmospheric Pressure : 14.2 psi					

CALIBRATION RESULTS

Sr. No.	Nominal Capacity (µl) at 27 °C	Actual Volume (µl) at 27 °C	Error (µl)	Expanded Uncertainty (±) µl
1	20	20.39	-0.39	0.13
2	100	100.94	-0.94	0.92
3	200	201.36	-1.36	0.92

End of Certificate

Hemraj
Calibrated By

Hemraj

Calibration Engineer



Reviewed/Approved By

L.M. Mahato

Chief Executive Officer (CEO)

- Note: -
- The calibration results reported in this certificate are valid at the time of and stated condition of measurement.
 - This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 - Total liability of our Organization is limited to the invoiced amount.
 - Calibration Certificate without signature are not valid.
 - All our certificates are metrological traceable.
 - The results relate only to the item calibrated.
 - UUC: Unit Under Calibration.
 - Expanded Uncertainty of measurement expressed at approximately the 95 % level of confidence, using a coverage factor of k = 2.
 - Scan this QR code to download our NABL Accreditation Scope & Certificate.



Working : W.Z. 1390/Z-2, Nangal Raya Extn., Pankha Road, New Delhi - 110046

Regd. Office : R.Z. 56A, Indra Park, Gali No. - 25, Palam Colony, New Delhi - 110045

M. : 9958445875, 9313768426, 9999761885 E - mail : suryaadrashcalibration@gmail.com



Scanned with OKEN Scanner