

## FLOW C



NABL Accredited Calibration Lab as per ISO/IEC 17025:2017

#71, Koorgalli Industrial Area, Mysuru-570 018. Mob: 98865 02708, Email: vinay1flowcal@gmail.com, www.flowcal.in

FCL/FM/CL/0

0001			ICATE OF CALIBRATION	457	
SRF No.	F0132	Certificate No.	FCL/24/F0132-04	ULR No.	CC310322000000732F
Date of SRF 0	03-09-2024	Date of Cal.	03-09-2024	Next Cal Due	02-09-2025
Issue Date 0	05-09-2024	Cal Procedure No.	FCL-SOP-MECH-02	Reference Std.	SANAS-TR 45-01
No of Pages 0	)1	Page No.	01 of 01	Status on Receipt	Satisfactory
SCIPLINE-GROUP: MECHANICAL - Acceleration and Speed			Cal Done At	Lab ,Flowcal	

## Customer Name & Full Address:

M/s. KM Doddi Governament Hospital

Manches Dietrict-571422

ICTC Community Health Center malavalli Maddur Street, K.M. Doddi. Bharathi Nagar, Maddur Talluk

	Manuya District-3	1422		
Nomenclature	- <i>897</i>	Digital Centrifuge	Model	R-8CBL
Make	Agents .	Remi	Range	6000 RPM
Serial. No	-4/4	6549	Resolution	10 RPM
ID No		GEN/RPM-01	Accuracy	≥ ± 0.05%
Location	7	GENERAL LAB	MON MARKET	The state of the s

STANDARDS USED FOR CALIBRATION AND TRACEABILITY DETAILS						
Nomenclature	Make/Model	SI. No / ID No.	Traceable to	Validity		
Digital Tachometer	Lutron & DT2236	Q652919/FCL-A&S-01	EQN/CRF/24080065	07-08-2025		

**Environmental Condition:** MODE :NON CONTACT

Temperature: 20.4

**Humidity:** 

Calibration Re	sults:	and of the villey be		All Readings are in RPM
SI.NO	DUC Readings	Standard readings	Observed Deviation	Computed Measurement Uncertainty (±) in %
1.7%	100.0	99.5	0.5	2.50
2	500.0	499.4	0.6	2.50
2	1000	998.6	1.4	2.50
4	2000	1997	3	2.50
5	3000	2994	and the table to the same of t	2.50
6	6000	5988	12	2.50

## Remarks:

- 1. The Measurement Uncertainty is estimated at a confidence level of 95.45% with a coverage factor k=2.
- 2. This Calibration Certificate refers only to the Particular item submitted for calibration. Calibration points given as per customer request.
- 3.NABL Accerdited Calibration Lab as per ISO/IEC 17025:2017 with vide Certificate No:CC-3103
- 4.The Usage of NABL symbol is as per NABL guidelines given in NABL 133.The Calibration Certificate relates only to the above DUC.
- 5.Calibration Certificate Shall not be reproduced except in full, without written approval of the Flowcal.
- Standard maintained are traceable to National / International Standard through accredited laboratories.

Calibrated by

Calibration Engineer

Authorized by

Vinay Kumar

CEO

\*\*\*\*\*End of Calibration Certificate\*\*\*\*

