



SURYAADARASH CALIBRATION AND CONSULTANT (OPC) PVT. LTD.

सूर्यआदर्श कैलिब्रेशन एंड कंसल्टेंट (ओपीसी) प्राइवेट लिमिटेड

Formerly UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY

NABL Accredited Calibration Laboratory As per ISO/IEC 17025:2017



CC - 3287

CALIBRATION CERTIFICATE

Format No.: SCCPL-QF-7.8.1	Certificate No.: SCCPLCC20240927-008	ULR No.: CC328724000013908F
Group: Volume		Discipline: Mechanical

Customer's Name & Address M/s. Wellcare Path Lab 80, Shree Balaji Enclave, Dhakoli, Zirakpur, Punjab-160104	Page No.: 1 of 1	Calibration Performed: At Lab
	Date of Receipt: 27-09-2024	
	Service Request No.: SCCPLSRN20240927-002	
	Certificate Issue Date: 30-09-2024	
	Date of Calibration: 28-09-2024	
Next Due Date (Suggested by Customer): 27-09-2025		

Equipment Description (UUC) & Identification	
Name of the Instrument: Micro pipette	Make: Dragon Lab
	Model: -----
Range: 10 µl to 100 µl	Sr. No.: YE239BG0020168
Least Count: 1 µl	I.D. No.: WPL/MP-03
Condition of Instrument: Ok	Location: -----

Master Equipment Used					
Sr. No.	Instrument's Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	Digital Weighing Balance	0 to 220 g & SCCPL/CAL/005	SCGPL	SCGPLCC20240113-002	14-01-2025
Environmental Conditions			Calibration Procedure	Calibration Reference Standard	
Temperature: 27 ± 3 °C, Humidity: 45 to 80 %RH			SCGPL 03 SOP 018	ISO 8655	
Atmospheric Pressure : 14.2 psi					

CALIBRATION RESULTS

Sr. No.	Nominal Capacity (µl) at 27 °C	Actual Volume (µl) at 27 °C	Error (µl)	Expanded Uncertainty (±) µl
1	10	10.39	-0.39	0.13
2	50	50.57	-0.57	0.13
3	100	100.86	-0.86	0.92

End of Certificate

Hemraj
Calibrated By
Hemraj

Calibration Engineer

L.M. Mahato
Reviewed/Approved By
L.M. Mahato

Chief Executive Officer (CEO)

- Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.
2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
3. Total liability of our Organization is limited to the invoiced amount.
4. Calibration Certificate without signature are not valid.
5. All our certificates are metrological traceable.
6. The results relate only to the item calibrated.
7. UUC: Unit Under Calibration.
8. Expanded Uncertainty of measurement expressed at approximately the 95 % level of confidence, using a coverage factor of k = 2.
9. Scan this QR code to download our NABL Accreditation Scope & Certificate.



Working : W.Z. 1390/Z-2, Nangal Raya Extn., Pankha Road, New Delhi - 110046

Regd. Office : R.Z. 56A, Indra Park, Gali No. - 25, Palam Colony, New Delhi - 110045
M. : 9958445875, 9313768426, 9999761885 E - mail : suryaadrashcalibration@gmail.com