



CALIBRATION CERTIFICATE

Format No.: SCCPL-QF-7.8.1		Certificate No.: SCCPLCC20240927-002		ULR No.: CC328724000013902F	
Group: Acceleration & Speed			Discipline: Mechanical		
Customer's Name & Address			Page No.: 1 of 1	Calibration Performed: At Site	
M/s. Wellcare Path Lab 80, Shree Balaji Enclave, Dhakoli, Zirakpur, Punjab-160104			Date of Receipt: 27-09-2024		
			Service Request No.: SCCPLSRN20240927-002		
			Certificate Issue Date: 30-09-2024		
			Date of Calibration: 28-09-2024		
Next Due Date (Suggested by Customer): 27-09-2025					
Equipment Description (UUC) & Identification					
Name of the Instrument: Digital RPM Indicator (Centrifuge Machine)			Make: Laby		
			Model: M-6		
Range: Up to 6000 RPM			Sr. No.: 2024-25/016		
Least Count: 5 RPM			I.D. No.: WPL/CF-01		
Condition of Instrument: Ok			Location: -----		
Master Equipment Used					
Sr. No.	Instrument's Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	Digital Tachometer	2.5 To 99999 RPM & SCCPL/CAL/014	Asian Technology	AT24000000170	25-02-2025
Environmental Conditions			Calibration Procedure	Calibration Reference Standard	
Temperature: 15 to 45 °C, Humidity: 30 to 80 %RH			SCCPL 03 SOP 033	----	

CALIBRATION RESULTS

Sr. No.	Set UUC Value (RPM)	Observed Ref. Standard Value (RPM)	Expanded Uncertainty (±)
1	500	498	4.1 % of rdg.
2	1000	992	4.1 % of rdg.
3	2000	1989	4.1 % of rdg.
4	4000	3982	4.1 % of rdg.
5	4500	4475	4.1 % of rdg.

End of Certificate

Amit
Calibrated By
Amit Kumar

Calibration Engineer

L.M. Mahato
Reviewed/Approved By
L.M. Mahato

Chief Executive Officer (CEO)

- Note:** - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.
 2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 3. Total liability of our Organization is limited to the invoiced amount.
 4. Calibration Certificate without signature are not valid.
 5. All our certificates are metrological traceable.
 6. The results relate only to the item calibrated.
 7. UUC: Unit Under Calibration.
 8. Expanded Uncertainty of measurement expressed at approximately the 95 % level of confidence, using a coverage factor of k = 2.
 9. Scan this QR code to download our NABL Accreditation Scope & Certificate.



Working : W.Z. 1390/Z-2, Nangal Raya Extn., Pankha Road, New Delhi - 110046

Regd. Office : R.Z. 56A, Indra Park, Gali No. - 25, Palam Colony, New Delhi - 110045
M. : 9958445875, 9313768426, 9999761885 E - mail : suryaadrashcalibration@gmail.com