

FL()W



NABL Accredited Calibration Lab as per ISO/IEC 17025:2017

#71, Koorgalli Industrial Area, Mysuru-570 018. Mob: 98865 02708, Email: vinay1flowcal@gmail.com, www.flowcal.in

FCL/FM/CL/06

		CE	RTIFICATE OF CALIE	BRATION	
	1		FCL/24/F0133-01	ULR No.	CC310324000000733F
RF No.	F0133	Certificate No.		Next Cal Due	02-09-2025
Date of SRF	03-09-2024	Date of Cal.	03-09-2024		DKD-R 5-1
ssue Date	05-09-2024	Cal Procedure No.	FCL-SOP-THE-01	Reference Std.	DKD-R 5-1
		D No.	01 of 01	Status on Receipt	Satisfactory
No of Pages	01	Page No.	01 01 01		Onsite
DISCIPI INF-GE	ROUP : Thermal-T	emperature		Cal Done At	Offsite
DISOIL FINE-OL	1001 1 1110111101				

Customer Name & Full Address:

M/s. kiragavalu Governament Hospital

CHC kiragavalu

Malavalli talluk, Mandya District

571424

57142	DUC DE	TAILS	
	ILR Refrigerator	Model / Type	GVR 225 AC
Nomenclature		Range	2 to 8 °C
Make	Godrej	Resolution	0.1 °C
Serial. No	NA		NA
ID No	ICTC//ILR-01	Accuracy	IVA
Location	ICTC LAB		

STANDARDS USED FOR CALIBRATION AND TRACEABILITY DETAILS

517	ANDARDS USED TO	N OALIBITATION THE		1 10 110
Nomenclature	Make/Model	SI. No / ID No.	Traceable to	Validity
4 Wire RTD Sensor With DMM	Tempsens & Yokogawa / PT - 100X1, Simplex & CA71	2113016632 & 2267021/FCL-TH-02 & FCL- TH-02A	TMS/24/64-01	15-03-2025

Temperature Bath Calibrator used as Source

Environmental Condition:

25.4 Temperature:

Humidity:

%RH 52

All Readings are in °C

Calibration Results:

Calibration F	Results:					
SI No	Set Temp	STD Readings	DUC Readings	Observed Deviation	Computed Measurement Uncertainty (±)	
				-0.371	0.8	
1	5	5.371	5.0	-0.371		

Remarks:

- 1. The Measurement Uncertainty is estimated at a confidence level of 95.45% with a coverage factor k=2.
- 2. This Calibration Certificate refers only to the Particular item submitted for calibration. Calibration points given as per customer request.
- 3.NABL Accerdited Calibration Lab as per ISO/IEC 17025:2017 with vide Certificate No:CC-3103
- 4. The Usage of NABL symbol is as per NABL guidelines given in NABL 133. The Calibration Certificate relates only to the above DUC.
- 5. Calibration Certificate Shall not be reproduced except in full, without written approval of the Flowcal.
- 6. Standard maintained are traceable to National / International Standard through accredited laboratories.

Calibrated by

(Calibration Engineer)

Authorized by

Vinay Kumar (CEO)

*****End of Calibration Certificate*****

