

# Parashar Micro Measurement Pvt. Ltd.



The way to Technology  
(21 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307  
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com  
Website : www.pmmpl.in, www.parashar.info



CC - 2795


## Calibration Certificate

ULR - CC279524000060819F		Page 1 of 1				
Certificate Number : PMM/180924/25-1		Date of Request	18.09.2024			
Calibrated For	M/s. Pathology Lab.	Calibrated on	18.09.2024			
	T.B. Combined Hospital, Lucknow.	Suggested Due Date	17.09.2025			
		Cert. Issue Date	25.09.2024			
<b>Description of Equipments</b>						
Nomenclature	Centrifuge	Accuracy/Class	Not Mentioned			
Range	3000 rpm	Condition of UUC	Physically Ok			
Resolution / Least Count	-----	Serial No.	-----			
Make & Model No.	Lab / T-3BLD	Party ID Mark No.	-----			
Material / Type	-----	Location	Pathology Lab			
<b>Master Equipment / Standard Used</b>						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Metravi	210813911 PMM/TM/09	PMMPL	PMM/1529	01.12.2024
Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration						
Calibration Performed at		Site	Environmental Condition(s)	Temperature = 23 ± 5° C Humidity = 30 to 75 % RH		
Calibration Procedure No.		CP/M/MV/S/03				
Reference Standard / Guideline		-----				
<b>Mechanical Calibration</b>						
<b>CALIBRATION RESULTS</b>						
Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)	Error (in rpm)	± Uncertainty in Measurement (in rpm)		
1	3000	3005	-5	13		

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor  $k = 2$ )

  
Calibrated By : Aakash Kumar  
Designation : Calibration Engineer



  
Approved By : Jai Prakash  
Designation : Tech. Manager / Auth. Sign.

Conditions : 1. This certificate refers only to the particular item submitted for calibration.

2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.

3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.