

# Parashar Micro Measurement Pvt. Ltd.



B-59, SECTOR-64, NOIDA (U.P.) 201 307  
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com  
Website : www.pmmpl.in, www.parashar.info



## Calibration Certificate

<b>ULR - CC279524000063623F</b>		<b>Page 1 of 1</b>				
<b>Certificate Number : PMM/051024/15-1</b>		<b>Date of Request</b>	05.10.2024			
<b>Calibrated For</b>	M/s. Clinical Pathology Sadar Hospital Araria, Bihar	<b>Calibrated on</b>	05.10.2024			
		<b>Suggested Due Date</b>	04.10.2025			
		<b>Cert. Issue Date</b>	08.10.2024			
<b>Description of Equipments</b>						
<b>Nomenclature</b>	Centrifuge	<b>Accuracy/Class</b>	Not Mentioned			
<b>Range</b>	0 - 5000 rpm	<b>Condition of UUC</b>	Physically Ok			
<b>Resolution / Least Count</b>	10 rpm	<b>Serial No.</b>	8622903			
<b>Make &amp; Model No.</b>	Bio Lab	<b>Party ID Mark No.</b>	-----			
<b>Material / Type</b>	-----	<b>Location</b>	-----			
<b>Master Equipment / Standard Used</b>						
<b>Sr.No.</b>	<b>Nomenclature</b>	<b>Make &amp; Model</b>	<b>Serial No. / I/D Mark</b>	<b>Calibrated From</b>	<b>Certificate No.</b>	<b>Due Date of Calibration</b>
1	Dig. Tachometer	Metravi	210813911 PMM/TM/09	PMMP, Noida	PMM/1529	01.12.2024
Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration						
<b>Calibration Performed at</b>		<b>Site</b>	<b>Environmental Condition(s)</b>	<b>Temperature = 23 ± 5° C</b>		
<b>Calibration Procedure No.</b>		CP/M/MV/S/03		<b>Humidity = 30 to 75 % RH</b>		
<b>Reference Standard / Guideline</b>		-----				
<b>Mechanical Calibration</b>						
<b>CALIBRATION RESULTS</b>						
<b>Sr.No.</b>	<b>Nominal Value in UUC (in rpm)</b>	<b>Actual Observed Value in Std. (in rpm)</b>	<b>Error (in rpm)</b>	<b>± Uncertainty in Measurement (in rpm)</b>		
1	1000	1001	-1	13		
2	2000	2002	-2	13		
3	3000	3005	-5	13		
4	4000	4006	-6	13		
5	5000	5007	-7	13		

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor  $k = 2$ )

Calibrated By : Vipin Kumar  
Designation : Calibration Engineer



Approved By : Jai Prakash  
Designation : Tech. Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.  
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.  
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.