



SRI CALIBRATIONS SERVICES

NABL ACCREDITED LABORATORY

12-10-335/3/A, S1 & S2, Nomula Laxmi Residency, Seetaphalmandi, Secunderabad - 500 061.



CC-2338

"BE ACCURATE & BE SAFE"

CALIBRATION CERTIFICATE

SCS/FQ-014

ULR-CC233824000020227F

Job No. : SCS/S-2549	Page : 1 of 2
Certificate No. : SCS/10-24/2549-03	Date of issue : 18-06-2024
Date of Calibration : 16-10-2024	Cal.Due Date : 15-10-2025

Customer Details : M/s.Star Diagnostics., Opp Sundharaiah Bhavan, Sundharaiah Bhavan Road, Ongole-523002.	Calibrated at. : Site Received Date. : 16-10-2024 Received Cond. : Good
---	---

Details of UUC:	
Description : Centrifuge	Sr.No. : 202005201-12
Make : Euro Diagnostic System	
Range : 0 to 4000 RPM & 0 to 99 mins	
Resolution : 10 RPM & 1min	

Details of Standard Instruments Used :			
Name	ID.No	Valid up to	Certificate No.
1.Digital Tachometer	SCS/P/29	04-11-2024	SCS/11-23/2516-02
2.Pro- Digit Timer	SCS/E/07	07-12-2024	SCS/12-23/2820-03

Procedure : WI-SP04,WI-SE07	Ref. Std. : SANAS-TR-45-01, MSL Technical Guide 8
-----------------------------	---

Environmental Details : Temperature : 30°C ± 5°C	Relative Humidity : 60 % ± 15% RH
---	-----------------------------------

CALIBRATION RESULTS

FUNCTION TESTED	STANDARD VALUE	UUC VALUE	ERROR	EXPANDED UNCERTAINTY(±)
RPM	RPM	RPM	RPM	RPM
	500.2	500	-0.2	3.4
	1001	1000	-1.0	3.4
	1501	1500	-1	3.4
	2002	2000	-2	3.4
	3002	3000	-2	3.4
	4003	4000	-3	3.4
TIME	min	min	min	sec
	5.02	5	-0.02	0.23
	30.03	30	-0.03	0.39
	60.03	60	-0.03	0.39
	90.04	90	-0.04	1.75

Calibrated by

N.Sridhar

Calibration Engineer



Approved by

A.Nageswara Rao

Managing Partner

