

**ACCURATE MEASUREMENTS, PLOT NO. 103, D BLOCK, MEETHAPUR EXTENSION, BADARPUR, NEW DELHI - 110044**

- NOTE :
- The calibration result reported in this certificate are related to item(s) calibrated & is valid at the time of and under stated condition of measurement.
  - This certificate should not be reproduced, except in full, without our prior permission in writing.
  - Calibration done by the laboratory is meant for scientific and industrial purpose only.

Calibrated By : - (Mohd. Yaqoob, Dy. TM)

Approved By : (S. Kaushik, QM)



REMARKS :-

Sr. No.	TEMPERATURE (in °C) @ ≈ 50 %RH	RELATIVE HUMIDITY (in %RH) @ ≈ 25°C
1	10.19	21.59
2	19.22	39.22
3	28.85	61.84
4	39.14	75.46
5	48.96	91.04
1	0.11	0.41
2	0.18	0.78
3	0.25	1.16
4	0.36	1.54
5	0.44	1.96
1	0.50	1.78
2	0.50	1.79
3	0.51	1.81
4	0.51	1.82
5	0.52	1.85

OBSERVATIONS

I. THERMAL CALIBRATION

Sr. No.	ITEM No.	DESCRIPTION	CAL. DUE DATE	CERTIFICATE No.	FARE LABS PVT. LTD.
1	AM/THM-001	DIG. THERMO HYGROMETER WITH SENSOR	01/06/2025	FL/C/TH/01062024-C007	

MASTERS(S) USED FOR CALIBRATION :-

STANDARDS USED FOR CALIBRATION ARE TRACEABLE TO NATIONAL OR INTERNATIONAL STANDARDS THROUGH NABL ACCREDITED LABORATORY.

TRACEABILITY :-

ITEM DESCRIPTION	Sr. No./ID No.	MAKE /MODEL	RANGE	LEAST COUNT	LOCATION
DIGITAL THERMO HYGROMETER	RT-25324/RT-25324	R-TEK	-50 to 70 °C & 10 - 99% RH	0.1 °C & 1% RH	MURSHIDABAD - 742401
CALIB. PERFORMED AT	LAB				
CALIBRATION PROCEDURE	AM-WI-121				
CONDITIONS	HUMIDITY (30 to 75) % RH				
ENVIRONMENTAL	TEMPERATURE (25 ± 4) °C				
ULR No.	CC228624000004913F				
CERTIFICATE No.	AM24CC04913				
CUSTOMER NAME & ADDRESS	M/s SALAR RURAL HOSPITAL	SALAR RH, BHARATPUR-II,			
		MURSHIDABAD - 742401			
RECEIPT DATE	17/08/2024				
CALIBRATION DATE	19/08/2024				
SUGG. DUE DATE	18/08/2025				
ISSUE DATE	21/08/2024				
CONDITION ON RECEIPT	SATISFACTORY				

ACCURATE MEASUREMENTS

CERT No. AM24CC04913

ID No. RT-25324

CAL DUE DT. 19/08/2024

SUGG DUE DT. 18/08/2025

9811826626

CC - 2286

**CALIBRATION CERTIFICATE**

