

CALIBRATED
 ACCURATE MEASUREMENTS
 CERT No. AM24CC04561
 ID No. 180040007
 CAL Dt. 02.08.2024
 SUGG DUE Dt. 01.08.2025
 9811826626
 MEASUREMENTS



CALIBRATION CERTIFICATE



AM-F-12

PAGE 1 OF 1

CC - 2286

CUSTOMER NAME & ADDRESS	: M/s MOGRAHAT RURAL HOSPITAL MOGRAHAT BLOCK - 2, PIN - 743355, SOUTH 24 PARGANAS	CERTIFICATE No.	: AM24CC04561
		ULR No.	: CC228624000004561F
		ENVIRONMENTAL CONDITIONS	: TEMPERATURE (20 ± 2)°C HUMIDITY (60 ± 20) % RH
		CALIB. PROCEDURE	: AM-WI-57M
		CALIB. PERFORMED AT	: LAB
		RECEIPT DATE	: 01/08/2024
		CALIBRATION DATE	: 02/08/2024
		SUGG. DUE DATE	: 01/08/2025
		ISSUE DATE	: 05/08/2024
		CONDITION ON RECEIPT	: SATISFACTORY

ITEM DESCRIPTION : **CENTRIFUGE**
 Sr. No. / ID No. : 180040007/180040007
 MAKE / MODEL : REMI/R-8C
 RANGE / SIZE : 5000 RPM
 LEAST COUNT : REFER OBSERVATION
 LOCATION : ---

TRACEABILITY :-
 STANDARDS USED FOR CALIBRATION ARE TRACEABLE TO NATIONAL OR INTERNATIONAL STANDARDS THROUGH NABL ACCREDITED LABORATORY.

MASTER(S) USED FOR CALIBRATION :-

Sr. No.	ITEM No.	DESCRIPTION	CALIB. DUE DATE	CERTIFICATE No.	CALIBRATED FROM
1	H.364908 / AM/TM 001	DIGITAL TACHOMETER	28/06/2025	AM24CC03721	ACCURATE MEASUREMENTS, DELHI


I. MECHANICAL CALIBRATION

OBSERVATIONS

Sr. No.	STANDARD (in RPM)	UNIT UNDER CALIBRATION (in RPM)	ERROR (in RPM)	Uncertainty at 95.45% C.L, k=2 (in RPM)
---------	-------------------	---------------------------------	----------------	---

NON CONTACT

1	4998	5000	2	4.2
---	------	------	---	-----



REMARK :-

END OF CERTIFICATE

Calibrated By: - (Vikas Sharma, Calib. Engg.)

Approved By: (S.Kaushik, QM)

NOTE :- 1 The calibration result reported in this certificate are related to item (s) calibrated & is valid at the time of and under stated condition of measurement.

2 This certificate should not be reproduced, expect in full, without our prior permission in writing.