



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029

*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

**EQAP CODE No. :** 64

**Distribution No.:** 165-B

**Month/Year:** September/2024

**Instrument ID:** Shenzhen  
Mindray Bio-Medical Electronics  
co.,Ltd.

**Model Name.:** Auto  
Hematology Analyzer BC-6200

**Serial No.:** TW-91000342

**Name & Contact No. of PT Co-ordinator:** Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi,  
Tel: 9013085730 , E-Mail : info@ishtmaimseqap.com

**Date of issue & status of the report:** 18-10-2024 [Final]

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	4.74	4.72	9.46	9.66	0.020	-0.33	0.02	0.1	0.005	-0.98
RBC x10 <sup>6</sup> /µl	1	4.39	4.36	8.75	8.4	0.008	1.63	0.03	0.03	0.002	0.00
Hb g/dl	1	12.9	12.8	25.7	25.2	0.019	1.12	0.1	0.1	0.007	0.00
HCT%	1	42.3	41.8	84.1	76.2	0.118	2.22	0.5	0.3	0.020	0.54
MCV-fl	1	96.3	95.9	192.2	182	0.239	1.35	0.4	0.3	0.020	0.27
MCH-Pg	1	29.3	29.3	58.6	60	0.047	-1.00	0	0.2	0.012	-0.90
MCHC-g/dl	1	30.5	30.4	60.9	66.3	0.103	-1.66	0.1	0.3	0.020	-0.67
Plt. x10 <sup>3</sup> /µl	1	158	153	311	342	1.301	-0.80	5	5	0.278	0.00
Retic %	2	1.2	0.8	2	7.5	0.188	-0.73	0.4	0.25	0.020	0.52

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
<b>DLC%</b>	3	Nrbcs=5 , Poly=7 L=78, E=12, Mono/Promono=3 , B1=0 P.M.=0, Mye=0, Meta=0, Other=
<b>RBC Morphology</b>	3	Predominantly Normocytic Normochromic, Mild hypochromia
<b>Diagnosis</b>	3	CHRONIC LYMPHOCYTIC LEUKEMIA
		Chronic Lymphoproliferative Disorder (CLPD)

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 165--B	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	382	379	85.49	93.67	2.11	4.49	12.40	1.84
<b>RBC x10<sup>6</sup>/µl</b>	1	382	382	89.01	87.17	3.4	6.28	7.59	6.55
<b>Hb g/dl</b>	1	382	382	80.89	90.58	8.9	4.45	10.21	4.97
<b>HCT%</b>	1	382	379	91.56	88.13	5.28	7.12	3.16	4.75
<b>MCV-fl</b>	1	382	377	95.76	85.15	3.45	7.96	0.79	6.89
<b>MCH-Pg</b>	1	382	378	87.83	76.46	6.08	18.52	6.09	5.02
<b>MCHC-g/dl</b>	1	382	378	94.97	89.95	2.38	4.76	2.65	5.29
<b>Plt. x10<sup>3</sup>/µl</b>	1	382	380	92.63	93.95	5.26	3.95	2.11	2.1
<b>ReticCount%</b>	2	382	347	96.54	81.84	2.59	1.15	0.87	17.01
<b>PS Assessment</b>	3	382	346	Satisfactory :93.99%, Borderline Sat. :3.14%, Unsatisfactory :2.87%					

**\*Comments:**

- 1). **Among Lab (EQA) : Results acceptable.**
- 2). **Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between " $0$  to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ( $0.3 \times \text{SDPA}$ ). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x} - \bar{y}$ ) should be smaller than the check value ( $0.3 \times \text{SDPA}$ ).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----