

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Urea	Erba Chem 7	11.00	mg/dl	-5.45
# HDL	Erba Chem 7	20.00	mg/dl	0
# ALT	Erba Chem 7	6.00	IU/L	0
# LDH	Erba Chem 7	42.00	U/L	0
X Sodium	Acculab Enlite	131.00	mmol/L	-6.41
! Phosphorous Inorganic	Erba Chem 7	3.40	mg/dl	-2.63
! Chloride	Acculab Enlite	96.00	mmol/L	-2.13
! Potassium	Erba Chem 7	4.43	mEq/L	-2.21
<p>Legend @ : z score ≤ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.</p>				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____ **Dated:** _____

Instrument : Erba Chem 7

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
X Urea	mg/dl	11.00	22.00	-5.45	-5.29
# HDL	mg/dl	20.00			-0.59
# ALT	U/L	6.00			2.72
# LDH	U/L	42.00			0.04
@ Glucose	mg/dl	86.00	88.07	-1.14	-0.07
@ Creatinine	mg/dl	4.15	3.91	0.54	1.02
@ Bilirubin Total	mg/dl	0.38	0.34	0.86	0.64
@ Bilirubin Direct	mg/dl	0.11	0.11	-0.43	-0.23
@ Total Protein	g/dL	6.50	6.56	-0.22	-0.37
@ Albumin	g/dL	3.90	3.78	-0.40	-0.33
@ Uric Acid	mg/dl	4.70	4.85	-0.19	-0.13
@ Calcium	mg/dl	8.70	9.13	0.57	0.76
! Phosphorous Inorganic	mg/dl	3.40	3.50	-2.63	-2.48
@ Cholesterol	mg/dl	156.00	154.45	0.51	0.77
@ Triglycerides	mg/dl	107.00	118.95	-1.72	0.60
@ LDL	mg/dl	107.00	--	0.55	1.18
@ AST	U/L	15.00	11.92	0.87	2.67
@ ALP	U/L	82.00	82.68	0.12	0.95
@ CK	U/L	69.00	63.75	0.78	1.62
@ Amylase	U/L	74.00	62.81	1.03	1.19
@ Lipase	U/L	40.00	49.33	-0.14	6.73
! Potassium	mmol/L	4.43	4.24	-2.21	-0.65

Instrument : Acculab Enlite

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
X Sodium	mmol/L	131.00	162.51	-6.41	-2.95
! Chloride	mmol/L	96.00	103.00	-2.13	-1.16

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	24
Not Evaluated Parameters	3
Evaluated Parameters	21
Outlier Parameters (X)	2
EQAS Score Biochemistry	90.48 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070