



Metropolis Healthcare Ltd., EQAS

SAI DIAGNOSTIC CENTRE, BOISAR (1332)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240207 and Sample No 01

Report Date : 18/10/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual


Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ Rheumatoid Factor Qualitative	IU/mL	Normal	Normal
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ HIV Ab	S/CO	Non-reactive	Non-reactive
@ Typhidot IgM	-	Negative	Negative
@ Widal test - Typhi O antigen (TO)	-	No Agglutination	No Agglutination
@ Widal test - Typhi H antigen (TH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination
@ ASO Qualitative	IU/mL	Normal	Normal
@ Dengue IgM	RU/ml	Negative	Negative
@ Dengue IgG	RU/ml	Negative	Negative

Instrument : Erba Chem 7

Analyte	Standard Unit	Result Value	Accepted Value
@ C-Reactive Protein Qualitative	mg/L	Abnormal	Abnormal

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	15
Not Evaluated Parameters	0
Evaluated Parameters	15
Outlier Parameters (X)	0
EQAS Score Immunoassay Extended	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070