



**KASTURBA MEDICAL COLLEGE**  
MANIPAL  
(A constituent unit of MAHE, Manipal)



**External Quality Assessment Scheme for HIV Serology**  
**Report of EQAS Proficiency Testing**

Date: 28/08/2024

YEAR: 2024-25

Round: 1

LAB ID: KA53ICTC05

Name of the participating laboratory: ICTC, GH Koppa, Chikkamagaluru District.

Date of Panel Distribution: 20/08/2024

Report of EQAS results obtained by your ICTC:

SAMPLE CODE	ICTC RESULT	SRL RESULT	REMARKS
A1	NEGATIVE	NEGATIVE	CONCORDANT
B2	NEGATIVE	NEGATIVE	CONCORDANT
C3	POSITIVE	POSITIVE	CONCORDANT
D4	NEGATIVE	NEGATIVE	CONCORDANT

You have scored – 100% (4/4).


Comments: Performance is satisfactory/  unsatisfactory.

Thanking You

Regards,

  
28/08/2024

Mrs.V.S.Anuradha  
Technical Officer  
State Reference Laboratory  
Department of Microbiology  
KMC, Manipal



Dr.Vandana K.E  
Prof & Head / SRL In-charge  
Department of Microbiology  
KMC, Manipal

**STATE REFERENCE LABORATORY**

**Department of Microbiology**

Kasturba Medical College, Manipal. 576104.

Karnataka, India. Ph: 91 820 2922717. e-mail : [srl.karnataka.kmc@gmail.com](mailto:srl.karnataka.kmc@gmail.com)



PROFICIENCY TESTING SCHEME FOR HIV SEROLOGY

KMC, Manipal, Karnataka  
KASTURBA MEDICAL COLLEGE, MANIPAL,  
DEPARTMENT OF MICROBIOLOGY, CENTER FOR BASIC SCIENCES, IIND FLOOR, 576104

Report released: 28-08-2024 12:18:37 PM

Proficiency Testing (PT) Report

Name of PT Panel Provider (NRL): NIMHANS, Bengaluru, Karnataka  
Name of Panel distributing Laboratory (SRL): KMC, Manipal, Karnataka

Participant Lab: M S D M GENERAL HOSPITAL KOPPA  
Laboratory ID: 2901143  
Round: First - 2024-25

Panel issued: 2024-08-20  
Panel received: 2024-08-20  
Panel tested: 2024-08-21


PARTICIPANT PERFORMANCE REPORT

Panel Member	Panel ID	Results of Participating Laboratory	Final Status	Laboratory Performance
Panel Member - 1	A1	Negative	Concordant	Satisfactory
Panel Member - 2	B2	Negative	Concordant	
Panel Member - 3	C3	Positive	Concordant	
Panel Member - 4	D4	Negative	Concordant	

This report is confidential and the performance is shared with the participating laboratory, respective National Reference Laboratory, Apex Laboratory and NACO only.  
Kindly contact SRL In-charge for any queries.

Report authorized by  
Dr. Vandana K.E- SRL In-Charge

-: End of report :-



Page 1 of 1

Administrative Medical Officer  
M.S.D.M. General Hospital  
KOPPA-577 126

This is a system generated report. No signature required