



FLOW CAL

TRUSTED AND RELIABLE CALIBRATION

NABL Accredited Calibration Lab as per ISO/IEC 17025:2017



CC-3103

#71, Koorgalli Industrial Area, Mysuru-570 018. Mob: 98865 02708, Email: vinay1flowcal@gmail.com, www.flowcal.in

FCL/FM/CL/06

CERTIFICATE OF CALIBRATION					
SRF No.	F0142	Certificate No.	FCL/24/F0142-01	ULR No.	CC310324000000797F
Date of SRF	27-09-2024	Date of Cal.	27-09-2024	Next Cal Due	26-09-2025
Issue Date	30-09-2024	Cal Procedure No.	FCL-SOP-THE-01	Reference Std.	DKD-R 5-1
No of Pages	01	Page No.	01 of 01	Status on Receipt	Satisfactory
DISCIPLINE-GROUP : Thermal-Temperature				Cal Done At	Onsite

Customer Name & Full Address:
M/s. ICTC GH,Koppa,Chikkamagaluru District.

DUC DETAILS			
Nomenclature	Digital Thermometer	Model / Type	---
Make	NA	Range	-50 to 150 °C
Serial. No	NA	Resolution	0.1 °C
ID No	ICTC/DTM-01	Accuracy	NA
Location	ICTC Lab		

STANDARDS USED FOR CALIBRATION AND TRACEABILITY DETAILS				
Nomenclature	Make/Model	SI. No / ID No.	Traceable to	Validity
4 Wire RTD Sensor With DMM	Tempsens & Yokogawa / PT - 100X1, Simplex & CA71	2113016632 & 2267021/FCL-TH-02 & FCL-TH-02A	TMS/24/64-01	15-03-2025

Temperature Bath Calibrator used as Source

Environmental Condition : Temperature : 25.4 °C Humidity: 52 %RH

Calibration Results: All Readings are in °C

SI No	Set Temp	STD Readings	DUC Readings	Observed Deviation	Computed Measurement Uncertainty (±)
1	-10	-9.978	-10.3	-0.322	0.8
2	0	0.063	0.2	0.137	0.8
3	30	30.084	30.3	0.216	0.8
4	60	60.273	60.5	0.227	0.8
5	90	90.185	90.4	0.215	0.8
6	120	120.123	120.3	0.177	0.8
7	150	150.154	150.5	0.346	0.8

- Remarks :**
- The Measurement Uncertainty is estimated at a confidence level of 95.45% with a coverage factor k=2.
 - This Calibration Certificate refers only to the Particular item submitted for calibration. Calibration points given as per customer request.
 - NABL Accredited Calibration Lab as per ISO/IEC 17025:2017 with vide Certificate No:CC-3103
 - The Usage of NABL symbol is as per NABL guidelines given in NABL 133. The Calibration Certificate relates only to the above DUC.
 - Calibration Certificate Shall not be reproduced except in full, without written approval of the Flowcal.
 - Standard maintained are traceable to National / International Standard through accredited laboratories.

Calibrated by

Sanju

(Calibration Engineer)



Authorized by

Vinay Kumar

(CEO)

*****End of Calibration Certificate*****