

HORIBA India Private Ltd.

246, Okhla Industrial Estate, Phase - III
110020, New Delhi, India
Tel: +91 (11) 4646 5000 / 4669 5001
Fax: +91 (11) 4669 5010 / 4646 5020

Date: 03/08/2024.


CALIBRATION CERTIFICATE

This is to certify that the Hematology analyzer HORIBA YUMIZEN H 550 bearing serial no: - 911YAXH02810 at Lupin Diagnostics center, Old Number 8, New No 19, Flowers Road, Purasaiwakkam, Chennai- 600084, was calibrated with CX496, Expiry of 05/08/2024, on **03/08/2024**.

The CV of repeatability check, carryover, calibrator and control values were all found within range.

New coefficients were calibrated and the instrument is working satisfactorily.
Next calibration is due on 02/08/2025

Calibration done by:
For HORIBA MEDICAL
HORIBA INDIA PVT LIMITED

A blue circular stamp with the text "HORIBA India Private Limited" around the perimeter and a star at the bottom. A handwritten signature in blue ink is written across the stamp.

V S Dhanesh kumar
Engineer- Customer support