



Alere Medical Pvt. Ltd.
No. 404, BPTP Park Centra,
NH - 8, Sector - 30
Opp. 32nd Milestone, Gurugram
Haryana 122001, India

T: +91 124 4569000
F: +91 124 4569011

Calibration Certificate

Instrument Name : URMETER-120
Manufacturer : Abbott Diagnostics Medical Pvt. Ltd..
Serial Number : 12212629
Customer : LUPIN DIAGNOSTICS, CHENNAI
MG TOWER, KILPAUK, CHENNAI TN -600084

Standard used : Self Calibration Mode

Standard Traceability Calibration : In House Certificate

The instrument function well in all the systems check, self check and maintenance Procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of 333 +10.

QC Performance : Acceptable

Self Calibration : Acceptable

Performance Qualification : Passed

Calibration Date . 09.07.2024

Calibration Passed

Calibration Valid up to 08.07.2025

For Alere Medical Pvt Ltd.

(Rohit R)
Engineer Technical Services
Mob:-09567271430

