



## PM & Calibration Certificate

Customer name	ESIC MODEL HOSPITAL
Address	BAPUNAGAR AHMEDABAD
Instrument	Auto Hematology Analyzer
Model	BC-6200
Serial No.	TW-91000342
Calibration Done on	11/06/2024
Next Calibration Due On or before	10/06/2025

**For Aspen Diagnostic Pvt. Ltd.**



Bhaumik Patel

Territory Manager Service



# CRITISURE MEDITECH

Technology meets Knowledge



CC-3498

## CALIBRATION CERTIFICATE

ULR Number:- CC349624000000171F

**Customer Name & Address:** M/s Chied Medical Officer, ESIC Model Hospital,  
Bapunagar, Ahmedabad, Gujarat-380024

### Detail of Device under Calibration

Equipment Name:- DOM.REFRIGERATER	Assess ID:- CM/CAL/ESIC-BPN/171
Make:- NA	Date of Calibration:- 13-08-2024
Model:- GL-185	Suggested due date:- 12-08-2025
Serial Number:- NA	Date of issue:- 18-08-2024
Location of Calibration:- on site-PATHOLOGY	Condition of Item:- Found Ok
Ambient Temperature:- 23.6°C	Note (if any)-
Cal. Procedure No.-SOP-16	Range:- Multi

### Master Equipment Details:

Sn	Standard Name	Make & Model	Serial Number	Certificate Number	Validity
1	Digital Thermometer	Nagman & 205	210908128	240409027	08-04-2025
2	Ele. Safety Analyzer	Fluke-ESA612	5544504	OHCC240146	10-04-2025
3					

## CALIBRATION RESULTS

Parameter name	UUC value	Limits	Result	Voltage (V)
Patient Leakage Current	2.3µA	<50 µA	PASS	L-N 227.9
Insulation Resistance	Open	200 mΩ	PASS	N-E 1.7
Body leakage Current	2.65µA	<100 µA	PASS	L-E 228.1
500V Applied test		PASS		

Parameter name	Standard reading	UUC Reading	Observation	Result
Temperature	°C	°C	°C	°C
	2	5	3	1.56
	5	10	5	2.42
	8	11	3	1.56



# CRITISURE MEDITECH

Technology meets Knowledge



ULR Number:- CC3496240000000171F

Parameter name	Standard reading	UUC Reading	Observation	Result

### Note:

- Entries in the Uncertainty column show the expanded measurement uncertainties evaluated and reported in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM) The coverage factor (k) is 2 with coverage probability of approximately 95.45%.
- The result reported is valid at the time of under the stated conditions of measurement.
- Any error in this report should be brought to our knowledge within 15 days from the date of this report.
- This certificate shall not be reproduced except in full or part without the written approval of the calibration facility.
- CRITISURE MEDITECH certifies that the performance of the above instrument has been verified using test equipment of known accuracy and had been calibrated using measurement standards traceable to the International System of Units (SI) through the National and international Standards.
- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of CRITISURE MEDITECH.

\*\*\*End of Certificate\*\*\*

Calibrated By

Omkar Dave  
(Technical Manager)



Authorized by

Vipul Bhatu  
(Quality Manager)