



Certificate of Conformity and Calibration

TO WHOMSOEVER IT MAY CONCERN

ISO 15189:2012 REQUIREMENTS REGARDING, "CALIBRATION AND VERIFICATION PROCEDURE"

All Transasia Bio-Medicals Ltd diagnostics products which are distributed and for which a certificate is issued are CE Marked. Transasia Bio-Medicals Ltd, Mumbai, manufacturers of diagnostic devices with company quality management system in compliance with standard ISO13485:2016 ;ISO9001:2015 .This means that all the processes in development and manufacturing of TBM products are guided by quality management system.

TBM declares and assure following

- The mentioned regulations require that production system and measuring devices are qualified and manufacturing and test procedures are validated as per ISO 13485:2016 and ISO 9001:2015 standard and it is assured through schedule maintenance and by regular qualification.
- TBM declares to have established procedure and to maintain it in order to assure the post marketing surveillance according to directive of 98/79/EC.
- All physical quantities, calibrators and controls used in TBM system are fully traceable to certified standards or reference materials.
- All TBM products are factory calibrated and final qc passed at the time of release.
- The performance of TBM system at customer site is assured if regular QC measurements, cleaning and maintenance procedure as described in the instruction for use or service documentation are performed
- Additional calibration or verification procedure is not required in order to assure the specified performances of every TBM system. Only if user deviates from the manufactures recommendation does he have to establish site specific calibration and verification procedure as part of his accreditation process.

Date:-26/04/2021

Manish Airan

Head Of Quality Department & Regulatory



CRITISURE MEDITECH

Technology meets Knowledge



CC-3496

CALIBRATION CERTIFICATE

ULR Number:- CC349624000000186F

Customer Name & Address: M/s Chied Medical Officer, ESIC Model Hospital,
Bapunagar,Ahmedabad, Gujarat-380024

Detail of Device under Calibration

Equipment Name:- CENTRIFUGE	Assess ID:- CM/CAL/ESIC-BPN/186
Make:- REMI	Date of Calibration:- 13-08-2024
Model:- R-8C PLUS	Suggested due date:- 12-08-2025
Serial Number:- ZGCN-07624	Date of issue:- 18-08-2024
Location of Calibration:- on site-PATHOLOGY LAB	Condition of Item:- Found Ok
Ambient Temperature:- 26.3°C	Note (if any)-
Cal. Procedure No.-SOP-43	Range:- Multi

Master Equipment Details:

Sn	Standard Name	Make & Model	Serial Number	Certificate Number	Validity
1	Techometer	Systems & HTM-890	19-013	240409026	08-04-2025
2	Ele. Safety Analyzer	Fluke-ESA612	5544504	OHCC240146	10-04-2025
3					

CALIBRATION RESULTS

Parameter name	UUC value	Limits	Result	Voltage (V)
Patient Leakage Current	18.2µA	<50 µA	PASS	L-N 228.7
Insulation Resistance	Open	200 mΩ	PASS	N-E 1.6
Body leakage Current	19.6µA	<100 µA	PASS	L-E 230.4
500V Applied test			PASS	

Parameter name	Standard reading	UUC Reading	Observation	Result
Rotation	RPM	RPM	RPM	RPM
	1000	1014	14	6.29
	2000	2015	15	6.72
	3000	3011	11	5



CRITISURE MEDITECH

Technology meets Knowledge



ULR Number:- CC3496240000000186F

Parameter name	Standard reading	UUC Reading	Observation	Result
Rotation	RPM	RPM	RPM	RPM
	4000	4024	24	10.59
	5000	5024	24	10.59
	6000	6021	21	9.3

Note:

- Entries in the Uncertainty column show the expanded measurement uncertainties evaluated and reported in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM) The coverage factor (k) is 2 with coverage probability of approximately 95.45%.
- The result reported is valid at the time of under the stated conditions of measurement.
- Any error in this report should be brought to our knowledge within 15 days from the date of this report.
- This certificate shall not be reproduced except in full or part without the written approval of the calibration facility.
- CRITISURE MEDITECH certifies that the performance of the above instrument has been verified using test equipment of known accuracy and had been calibrated using measurement standards traceable to the International System of Units (SI) through the National and international Standards.
- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of CRITISURE MEDITECH.

End of Certificate

Calibrated By

Omkar Dave
(Technical Manager)



Authorized by

Vipul Bhatu
(Quality Manager)



CRITISURE MEDITECH

Technology meets Knowledge



CALIBRATION CERTIFICATE

ULR Number:- CC3496240000000184F

Customer Name & Address: M/s Chied Medical Officer, ESIC Model Hospital,
Bapunagar,Ahmedabad, Gujarat-380024

Detail of Device under Calibration

Equipment Name:- DOM.REFRIGERATER	Assess ID:- CM/CAL/ESIC-BPN/184
Make:- WHIRPOOL	Date of Calibration:- 13-08-2024
Model:- IF 455D	Suggested due date:- 12-08-2025
Serial Number:- 2.1295E+15	Date of issue:- 18-08-2024
Location of Calibration:- on site-PATHOLOGY (GROUND FLOOR)	Condition of Item:- Found Ok
Ambient Temperature:- 23.6°C	Note (if any)-
Cal. Procedure No.-SOP-16	Range:- Multi

Master Equipment Details:

Sn	Standard Name	Make & Model	Serial Number	Certificate Number	Validity
1	Digital Thermometer	Nagman & 205	210908128	240409027	08-04-2025
2	Ele. Safety Analyzer	Fluke-ESA612	5544504	OHCC240146	10-04-2025
3					

CALIBRATION RESULTS

Parameter name	UUC value	Limits	Result	Voltage (V)
Patient Leakage Current	2.3µA	<50 µA	PASS	L-N 227.9
Insulation Resistance	Open	200 mΩ	PASS	N-E 1.7
Body leakage Current	2.65µA	<100 µA	PASS	L-E 228.1
500V Applied test		PASS		

Parameter name	Standard reading	UUC Reading	Observation	Result
Temperature	°C	°C	°C	°C
	2	5	3	1.56
	5	10	5	2.42
	8	10	2	1.13



CRITISURE MEDITECH

Technology meets Knowledge



ULR Number:- CC349624000000184F

Parameter name	Standard reading	UUC Reading	Observation	Result

Note:

- Entries in the Uncertainty column show the expanded measurement uncertainties evaluated and reported in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM) The coverage factor (k) is 2 with coverage probability of approximately 95.45%.
- The result reported is valid at the time of under the stated conditions of measurement.
- Any error in this report should be brought to our knowledge within 15 days from the date of this report.
- This certificate shall not be reproduced except in full or part without the written approval of the calibration facility.
- CRITISURE MEDITECH certifies that the performance of the above instrument has been verified using test equipment of known accuracy and had been calibrated using measurement standards traceable to the International System of Units (SI) through the National and international Standards.
- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of CRITISURE MEDITECH.

End of Certificate

Calibrated By

Omkar Dave
(Technical Manager)



Authorized by

Vipul Bhatu
(Quality Manager)



CRITISURE MEDITECH

Technology meets Knowledge



CALIBRATION CERTIFICATE

ULR Number:- CC349624000000223F

Customer Name & Address: M/s Chied Medical Officer, ESIC Model Hospital,
Bapunagar, Ahmedabad, Gujarat-380024

Detail of Device under Calibration

Equipment Name:- DOM.REFRIGERATER	Assess ID:- CM/CAL/ESIC-BPN/223
Make:- GODREJ	Date of Calibration:- 13-08-2024
Model:- RT EON	Suggested due date:- 12-08-2025
Serial Number:- NA	Date of issue:- 18-08-2024
Location of Calibration:- on site-PATHOLOGY (GROUND FLOOR)	Condition of Item:- Found Ok
Ambient Temperature:- 24.5°C	Note (if any)-
Cal. Procedure No.-SOP-16	Range:- Multi

Master Equipment Details:

Sn	Standard Name	Make & Model	Serial Number	Certificate Number	Validity
1	Digital Thermometer	Nagman & 205	210908128	240409027	08-04-2025
2	Ele. Safety Analyzer	Fluke-ESA612	5544504	OHCC240146	10-04-2025
3					

CALIBRATION RESULTS

Parameter name	UUC value	Limits	Result	Voltage (V)
Patient Leakage Current	2.3µA	<50 µA	PASS	L-N 227.9
Insulation Resistance	Open	200 mΩ	PASS	N-E 1.7
Body leakage Current	2.65µA	<100 µA	PASS	L-E 228.1
500V Applied test		PASS		

Parameter name	Standard reading	UUC Reading	Observation	Result
Temperature	°C	°C	°C	°C
	2	3	1	0.7
	5	5	0	0.27
	8	6	-2	-0.59



CRITISURE MEDITECH

Technology meets Knowledge



ULR Number:- CC349624000000223F

Parameter name	Standard reading	UUC Reading	Observation	Result

Note:

- Entries in the Uncertainty column show the expanded measurement uncertainties evaluated and reported in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM) The coverage factor (k) is 2 with coverage probability of approximately 95.45%.
- The result reported is valid at the time of under the stated conditions of measurement.
- Any error in this report should be brought to our knowledge within 15 days from the date of this report.
- This certificate shall not be reproduced except in full or part without the written approval of the calibration facility.
- CRITISURE MEDITECH certifies that the performance of the above instrument has been verified using test equipment of known accuracy and had been calibrated using measurement standards traceable to the International System of Units (SI) through the National and international Standards.
- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of CRITISURE MEDITECH.

End of Certificate

Calibrated By

Omkar Dave
(Technical Manager)



Authorized by

Vipul Bhatu
(Quality Manager)



CRITISURE MEDITECH

Technology meets Knowledge



CALIBRATION CERTIFICATE

ULR Number:- CC3496240000000180F

Customer Name & Address: M/s Chied Medical Officer, ESIC Model Hospital,
Bapunagar, Ahmedabad, Gujarat-380024

Detail of Device under Calibration

Equipment Name:- DOM.REFRIGERATER	Assess ID:- CM/CAL/ESIC-BPN/180
Make:- LG	Date of Calibration:- 13-08-2024
Model:- GL-D241AHAN	Suggested due date:- 12-08-2025
Serial Number:- H08530	Date of issue:- 18-08-2024
Location of Calibration:- on site-PATHOLOGY (GROUND FLOOR)	Condition of Item:- Found Ok
Ambient Temperature:- 24.5°C	Note (if any)-
Cal. Procedure No.-SOP-16	Range:- Multi

Master Equipment Details:

Sn	Standard Name	Make & Model	Serial Number	Certificate Number	Validity
1	Digital Thermometer	Nagman & 205	210908128	240409027	08-04-2025
2	Ele. Safety Analyzer	Fluke-ESA612	5544504	OHCC240146	10-04-2025
3					

CALIBRATION RESULTS

Parameter name	UUC value	Limits	Result	Voltage (V)
Patient Leakage Current	2.3µA	<50 µA	PASS	L-N 227.9
Insulation Resistance	Open	200 mΩ	PASS	N-E 1.7
Body leakage Current	2.65µA	<100 µA	PASS	L-E 228.1
500V Applied test		PASS		

Parameter name	Standard reading	UUC Reading	Observation	Result
Temperature	°C	°C	°C	°C
	2	3	1	0.7
	5	9	4	1.99
	8	9	1	0.7



CRITISURE MEDITECH

Technology meets Knowledge



ULR Number:- CC3496240000000180F

Parameter name	Standard reading	UUC Reading	Observation	Result

Note:

- Entries in the Uncertainty column show the expanded measurement uncertainties evaluated and reported in accordance with the ISO Guide to the Expression of Uncertainty in Measurement (GUM) The coverage factor (k) is 2 with coverage probability of approximately 95.45%.
- The result reported is valid at the time of under the stated conditions of measurement.
- Any error in this report should be brought to our knowledge within 15 days from the date of this report.
- This certificate shall not be reproduced except in full or part without the written approval of the calibration facility.
- CRITISURE MEDITECH certifies that the performance of the above instrument has been verified using test equipment of known accuracy and had been calibrated using measurement standards traceable to the International System of Units (SI) through the National and International Standards.
- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of CRITISURE MEDITECH.

End of Certificate

Calibrated By

Omkar Dave
(Technical Manager)



Authorized by

Vipul Bhatu
(Quality Manager)



GLOBAL

BIO MEDICAL SERVICES INDIA



CC - 3459

CALIBRATION CERTIFICATE

ULR - CC40412400000567F

Certificate Number : ESIC/P/12	Date of Receipt : 15.08.2024	Page no : 1 of 1
Service request no : ATPL/2024/09/0053	Calibration done at : LAB	
Service request date : 14.08.2024	Suggested Date Of Next Calibration (As Requested By Customer) : 15.02.2025	
Calibration date : 15.08.2024		
Date of issue : 16.08.2024		

DETAILS OF CUSTOMER :-	M/s. Chief Medical Officer, ESIC MODEL HOSPITAL Bapunagar, Ahmedabad Gujarat-380024
-------------------------------	----------------------------------------------------------------------------------------------

DETAILS OF UNIT UNDER CALIBRATION (UUC)	
Nomenclature : Micropipette	Range : 100-1000 µl
Make : MICROLIT RBO	Least Count : 0.1 µl
Model : NA	Visual Inspection : Ok
Serial No : SL716331	Accuracy/Acceptance Criteria : ----
ID Number :	Type : ----
Location : BIOCHEMSITRY	

STANDARD INSTRUMENTS USED FOR CALIBRATION					
S.No	Nomenclature	Serial No	Identification No	Certificate No	Calibration Valid Upto
01	Electronic Semi Micro Balance	15114327	GBSI/MEC/WM-01	TVCSPL 24/02/0229-01	15.02.2025

Environmental Conditions	Calibration Procedure : SOP/MEC/004
Temperature : 25°C ± 5°C	Reference Std : ISO 8655-6
Humidity : 35 to 85% RH	

CALIBRATION RESULTS

Discipline :-Mechanical (Volume)

S.No	UUC Set Value	Measured Volume @ 27°C	Error	Uncertainty (±)	PASS/FAIL
	µl	µl	µl	µl	
1	100	100.43	-0.43	0.16	PASS
2	300	301.50	-1.50	0.93	PASS
3	500	502.15	-2.15	0.93	PASS
4	700	702.50	-2.50	0.93	PASS
5	1000	1002.98	-2.98	0.93	PASS

* UUC Means Unit Under Calibration
 * Reported Measured Volume at Average of 10 Measurements

- Remarks :**
- This certificate refers only to the particular item submitted for calibration.
 - This certificate shall not be reproduced, except in full, without the written permission of Director of Global Biomedical Services .
 - Results reported are valid at the time of & under the stated conditions of measurement.
 - Laboratory Standards are Traceable to National and International Standards.
 - The reported expanded uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by coverage factor k=2, which is for a normal distribution corresponds to a coverage probability and level of confidence of approximately 95%.
- ***** End Of Certificate *****

Calibrated by
 Calibrated by

 Jyothi
 (Technical Manager/Calibration Engineer)



Authorised Signatory
 Authorised Signatory

 V. Naidu Babu Vanam
 (Technical Manager/Quality Manager)



GLOBAL

BIO MEDICAL SERVICES INDIA



CALIBRATION CERTIFICATE

ULR - CC40412400000592F

Certificate Number : ESICP/11	Date of Receipt : 15.08.2024	Page no : 1 of 1
Service request no : ATPL/2024/09/0053	Calibration done at : LAB	
Service request date : 14.08.2024	Suggested Date Of Next Calibration (As Requested By Customer) : 15.02.2025	
Calibration date : 15.08.2024		
Date of issue : 16.08.2024		

DETAILS OF CUSTOMER :-	M/s. Chief Medical Officer, ESIC Model Hospita; Bapunagar, Ahmedabad Gujarat-380024
-------------------------------	----------------------------------------------------------------------------------------------

DETAILS OF UNIT UNDER CALIBRATION (UUC)	
Nomenclature : Micropipette	Range : 2- 20 µl
Make : GLASSCO	Least Count : 0.1 µl
Model : NA	Visual Inspection : Ok
Serial No : SL716331	Accuracy/Acceptance Criteria : ----
ID Number :	Type : ----
Location : BIOCHEMISTRY	

STANDARD INSTRUMENTS USED FOR CALIBRATION					
S.No	Nomenclature	Serial No	Identification No	Certificate No	Calibration Valid Upto
01	Electronic Semi Micro Balance	15114327	GBSIMEC/WM-01	TVCSPL 24/02/0229-01	15.02.2025

Environmental Conditions		Calibration Procedure : SOP MEC/004	
Temperature : 25°C ± 5°C		Reference Std : ISO 8655-6	
Humidity : 35 to 85% RH			

CALIBRATION RESULTS

Discipline :-Mechanical (Volume)

UUC Set Value	Measured Volume @ 27°C	Error	Uncertainty (±)	PASS/FAIL
µl	µl	µl	µl	
20	20.11	-0.11	0.16	PASS

* UUC Means Unit Under Calibration

* Reported Measured Volume at Average of 10 Measurements


Remarks :


- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of Director of Global Biomedical services.
- Results reported are valid at the time of & under the stated conditions of measurement.
- Laboratory Standards are Traceable to National and International Standards.
- The reported expanded uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by coverage factor k=2, which is for a normal distribution corresponds to a coverage probability and level of confidence of approximately 95%.

***** End Of Certificate *****

Calibrated by

Authorised Signatory

Calibrated by

 Jyothi
 (Technical Manager)

Authorised Signatory

 (Technical Manager/Quality Manager)
 Naidu Babu Vanam
 (Quality Manager)



GLOBAL BIOMEDICAL SERVICES INDIA

10-35-5/1,3rd floor, Near Federal Bank, Bhanoji Nagar, Visakhapatnam, Andhra Pradesh-530002.

+91 9247871397 | +91 9494320719 | 0891-2941397, Email id: globalbiomedicalservices@gmail.com, www.gbsindia.co.in

WE ARE PROVIDING CALIBRATION SERVICES ALL OVER INDIA



GLOBAL

BIO MEDICAL SERVICES INDIA



CALIBRATION CERTIFICATE

ULR - CC40412400000583F

Certificate Number : ESIC/P/10	Date of Receipt : 15.08.2024	Page no : 1 of 1
Service request no : ATPL/2024/09/0053	Calibration done at : LAB	
Service request date : 14.08.2024	Suggested Date Of Next Calibration (As Requested By Customer) : 15.02.2025	
Calibration date : 15.08.2024		
Date of issue : 16.08.2024		

DETAILS OF CUSTOMER :-	M/s. Chief Medical Officer, ESIC Model Hospital, Bapunagar, Ahmedabad Gujarat-380024
-------------------------------	-----------------------------------------------------------------------------------------------

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Nomenclature : Micropipette	Range : Fixed 5 µl
Make : PFACT	Least Count : ---
Model : NA	Visual Inspection : Ok
Serial No : QE595302	Accuracy/Acceptance Criteria : ---
ID Number :	Type : ---
Location : BIOCHEMISTRY	

STANDARD INSTRUMENTS USED FOR CALIBRATION

S.No	Nomenclature	Serial No	Identification No	Certificate No	Calibration Valid Upto
01	Electronic Semi Micro Balance	15114327	GBSI/MEC/WM-01	TVCSP 24/02 0229-01	15.02.2025

Environmental Conditions	Calibration Procedure : SOP/MEC/004
Temperature : 25°C ± 5°C	Reference Std : ISO 8655-6
Humidity : 35 to 85% RH	

CALIBRATION RESULTS

Discipline :-Mechanical (Volume)

S.No	UUC Set Value	Measured Volume @ 27°C	Error	Uncertainty (±)	PASS/FAIL
	µl	µl	µl	µl	
1.00	5.00	5.02	-0.2	0.93	PASS

* UUC Means Unit Under Calibration

* Reported Measured Volume at Average of 10 Measurements


Remarks :

- This certificate shall not be reproduced, except in full, without the written permission of Director of Global Biomedical Services.
- Results reported are valid at the time of & under the stated conditions of measurement.
- Laboratory Standards are Traceable to National and International Standards.
- The reported expanded uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by coverage factor k=2, which is for a normal distribution corresponds to a coverage probability and level of confidence of approximately 95%.


***** End Of Certificate *****

Calibrated by

Authorised Signatory

Calibrated by

 Jayanthi Engineer
 Jayanthi
 (Technical Manager)



Authorised Signatory

 (Technical Manager/Quality Manager)
 Nandha Babu Varani
 (Quality Manager)

GLOBAL BIOMEDICAL SERVICES INDIA

10-35-5-1,3rd floor, Near Federal Bank, Bharoju Nagar, Visakhapatnam, Andhra Pradesh-530002

+91 9247871397 | +91 9494320719 | 0891-2941397, Email id: globalbiomedicalservices@gmail.com, www.gbsindia.co.in

WE ARE PROVIDING CALIBRATION SERVICES ALL OVER INDIA



CALIBRATION CERTIFICATE

ULR - CC40412400000589F

Certificate Number : ESIC/P/9	Date of Receipt : 15.08.2024	Page no : 1 of 1
Service request no : ATPL/2024/09/0053	Calibration done at : LAB	
Service request date : 14.08.2024	Suggested Date Of Next Calibration (As Requested By Customer) : 15.02.2025	
Calibration date : 15.08.2024		
Date of issue : 16.08.2024		

DETAILS OF CUSTOMER :-	M/s. Chief Medical Officer, ESIC Model Hospital Bapunagar, Ahmedabad Gujarat-380024
-------------------------------	----------------------------------------------------------------------------------------------

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Nomenclature : Fixed Micropipette	Range : Fixed 1000 µl
Make : LABQUEST	Least Count : ----
Model : NA	Visual Inspection : OK
Serial No : SF696102	Accuracy/Acceptance Criteria : ----
ID Number : PDU251	Type : ----
Location : BIOCHEMISTRY	

STANDARD INSTRUMENTS USED FOR CALIBRATION

S.No	Nomenclature	Serial No	Identification No	Certificate No	Calibration Valid Upto
01	Electronic Semi Micro Balance	15114327	GBSIMEC/WM-01	TVCSP 24/02/0229-01	15.02.2025

Environmental Conditions	Calibration Procedure : SOP/MEC/004
Temperature : 25 ± 5°C	Reference Std : ISO 8655-6
Humidity : 35 to 85 % RH	

CALIBRATION RESULTS

Discipline :-Mechanical (Volume)

UUC Set Value	Measured Volume @ 27°C	Error	Uncertainty (±)	PASS/FAIL
µl	µl	µl	µl	
1000	1003.95	-3.95	0.93	PASS

* UUC Means Unit Under Calibration

* Reported Measured Volume at Average of 10 Measurements

Remarks :

- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of Director of Global Biomedical Services
- Results reported are valid at the time of & under the stated conditions of measurement.
- Laboratory Standards are Traceable to National and International Standards.
- The reported expanded uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by coverage factor k=2, which is for a normal distribution corresponds to a coverage probability and level of confidence of approximately 95%.

***** End Of Certificate *****

Calibrated by

Calibrated by

Jyothi
(Technical Manager) Calibration Engineer



Authorised Signatory

Authorised Signatory

Naidu Babu Vanani
(Technical Manager) Quality Manager

GLOBAL BIOMEDICAL SERVICES INDIA

10-35-5/1,3rd floor, Near Federal Bank, Bhanoji Nagar, Visakhapatnam, Andhra Pradesh-530002.
+91 9247871397 | +91 9494320719 | 0891-2941397, Email id: globalbiomedicalservices@gmail.com. www.gbsindia.co.in
WE ARE PROVIDING CALIBRATION SERVICES ALL OVER INDIA



GLOBAL

BIO MEDICAL SERVICES INDIA



CALIBRATION CERTIFICATE

ULR - CC40412400000599F

Page no : 1 of 1

Certificate Number : ESIC/P/8	Date of Receipt : 15.08.2024
Service request no : ATPL/2024/10/0053	Calibration done at : LAB
Service request date : 14.08.2024	Suggested Date Of Next Calibration (As Requested By Customer) : 15.02.2025
Calibration date : 15.08.2024	
Date of issue : 16.08.2024	

DETAILS OF CUSTOMER :-	M/s. Chief Medical Officer, ESIC Model Hospital, Bapunagar, Ahmedabad Gujarat-380024
-------------------------------	-----------------------------------------------------------------------------------------------

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Nomenclature : Micropipette	Range : Fixed 20 µl
Make : PRECIA	Least Count : —
Model : NA	Visual Inspection : Ok
Serial No : NA	Accuracy/Acceptance Criteria : —
ID Number :	Type : —
Location : BIOCHEMISTRY	

STANDARD INSTRUMENTS USED FOR CALIBRATION

S.No	Nomenclature	Serial No	Identification No	Certificate No	Calibration Valid Upto
01	Electronic Semi Micro Balance	15114327	GBSI/MEC/WM-01	TVCSPL 24/02/0229-01	15.02.2025

Environmental Conditions	Calibration Procedure : SOP/MEC/004
Temperature : 25°C ± 5°C	Reference Std : ISO 8655-6
Humidity : 35 to 85% RH	

CALIBRATION RESULTS

Discipline :-Mechanical (Volume)

UUC Set Value	Measured Volume @ 27°C	Error	Uncertainty (±)	PASS/FAIL
µl	µl	µl	µl	
20	20.15	-0.15	0.22	PASS

- * UUC Means Unit Under Calibration
- * Reported Measured Volume at Average of 10 Measurements

Remarks :

- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of Director of Global Biomedical Services .
- Results reported are valid at the time of & under the stated conditions of measurement.
- Laboratory Standards are Traceable to National and International Standards.
- The reported expanded uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by coverage factor k=2, which is for a normal distribution corresponds to a coverage probability and level of confidence of approximately 95%.

***** End Of Certificate *****

Calibrated by

Authorised Signatory

Calibrated by
Calibration Engineer
Jyothi
Jyothi
(Technical Manager)



Authorised Signatory
(Technical Manager/Quality Manager)
V. Naidu Babu Vanam
Naidu Babu Vanam
(Quality Manager)

GLOBAL BIOMEDICAL SERVICES INDIA

10-35-5/1,3rd floor, Near Federal Bank, Bhanaji Nagar, Visakhapatnam, Andhra Pradesh-530002.

+91 9247871397 | +91 9494320719 | 0891-2941397, Email id: globalbiomedicalservices@gmail.com. www.gbsindia.co.in

WE ARE PROVIDING CALIBRATION SERVICES ALL OVER INDIA



GLOBAL

BIO MEDICAL SERVICES INDIA



CC - 3459

CALIBRATION CERTIFICATE

ULR - CC40412400000598F

Certificate Number : ESIC/P/7	Page no : 1 of 1
Service request no : ATPL/2024/10/0053	Date of Receipt : 15.08.2024
Service request date : 14.08.2024	Calibration done at : LAB
Calibration date : 15.08.2024	Suggested Date Of Next Calibration (As Requested By Customer) : 15.02.2025
Date of issue : 16.08.2024	

DETAILS OF CUSTOMER :-	M/s. Chief Medical Officer, ESIC Model Hospital Bapunagar- Ahmedabad Gujarat-380024
-------------------------------	----------------------------------------------------------------------------------------------

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Nomenclature : Micropipette	Range : FIXED 25 µl
Make : MICROLIT	Least Count : —
Model : RBO	Visual Inspection : Ok
Serial No : 20407116	Accuracy/Acceptance Criteria : —
ID Number : NA	Type : —
Location : BIOCHEMISTRY	

STANDARD INSTRUMENTS USED FOR CALIBRATION

S.No	Nomenclature	Serial No	Identification No	Certificate No	Calibration Valid Upto
01	Electronic Semi Micro Balance	15114327	GBSI/MEC/WM-01	TVCSP 24/02/0229-01	15.02.2025

Environmental Conditions	Calibration Procedure : SOP/MEC/004
Temperature : 25°C ± 5°C	Reference Std : ISO 8655-6
Humidity : 35 to 85% RH	

CALIBRATION RESULTS

Discipline :-Mechanical (Volume)

UUC Set Value	Measured Volume @ 27°C	Error	Uncertainty (±)	PASS/FAIL
µl	µl	µl	µl	
25	25.93	-0.93	0.93	PASS

• UUC Means Unit Under Calibration

• Reported Measured Volume at Average of 10 Measurements

Remarks :

- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of Director of Global Biomedical Services .
- Results reported are valid at the time of & under the stated conditions of measurement.
- Laboratory Standards are Traceable to National and International Standards.
- The reported expanded uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by coverage factor k=2, which is for a normal distribution corresponds to a coverage probability and level of confidence of approximately 95%.

***** End Of Certificate *****

Calibrated by

Authorised Signatory

Calibrated by
Calibration Engineer

Jyothi
(Technical Manager)



Authorised Signatory
(Technical Manager/Quality Manager)

Nandu Babu Vanam
(Quality Manager)

GLOBAL BIOMEDICAL SERVICES INDIA

10-35-5/1,3rd floor, Near Federal Bank, Bhanoji Nagar, Visakhapatnam, Andhra Pradesh-530002.

+91 9247871397 | +91 9494320719 | 0891-2941397, Email id: globalbiomedicalservices@gmail.com, www.gbsindia.co.in

WE ARE PROVIDING CALIBRATION SERVICES ALL OVER INDIA



GLOBAL

BIO MEDICAL SERVICES INDIA



CALIBRATION CERTIFICATE

ULR - CC40412400000590F

Certificate Number : ESIC/P/6	Page no : 1 of 1
Service request no : ATPL/2024/10/0053	Date of Receipt : 15.08.2024
Service request date : 14.08.2024	Calibration done at : LAB
Calibration date : 15.08.2024	Suggested Date Of Next Calibration (As Requested By Customer) : 15.02.2025
Date of issue : 16.08.2024	

DETAILS OF CUSTOMER :-	M/s. Chief Medical Officer ESIC Model Hospital bapunagar- Ahmedabad Gujarat-380024
-------------------------------	---------------------------------------------------------------------------------------------

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Nomenclature : Micropipette	Range : 100-1000 µl
Make : MICROLIT	Least Count : 1 µl
Model : RBO	Visual Inspection : Ok
Serial No : 23404917	Accuracy/Acceptance Criteria : ---
ID Number : NA	Type : ---
Location : BIOCHEMISTRY	

STANDARD INSTRUMENTS USED FOR CALIBRATION

S.No	Nomenclature	Serial No	Identification No	Certificate No	Calibration Valid Upto
01	Electronic Semi Micro Balance	15114327	GBSI/MEC/WM-01	TVCSP 24/02/0229-01	15.02.2025

Environmental Conditions	Calibration Procedure : SOP/MEC/004
Temperature : 25°C ± 5°C	Reference Std : ISO 8655-6
Humidity : 35 to 85% RH	

CALIBRATION RESULTS

Discipline :-Mechanical (Volume)

S.No	UUC Set Value	Measured Volume @ 27°C	Error	Uncertainty (±)	PASS/FAIL
	µl	µl	µl	µl	
1	100	100.36	-0.36	0.16	PASS
2	300	301.10	-1.10	0.93	PASS
3	500	501.85	-1.85	0.93	PASS
4	700	702.64	-2.64	0.93	PASS
5	1000	1002.98	-2.98	0.93	PASS

- * UUC Means Unit Under Calibration
- * Reported Measured Volume at Average of 10 Measurements

Remarks :

- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of Director of Global Biomedical Services.
- Results reported are valid at the time of & under the stated conditions of measurement.
- Laboratory Standards are Traceable to National and International Standards.
- The reported expanded uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by coverage factor k=2, which is for a normal distribution corresponds to a coverage probability and level of confidence of approximately 95%.

***** End Of Certificate *****

Calibrated by

Calibrated by

(Signature)
S. Jayothi
(Technical Manager/Engineer)



Authorised Signatory

Authorised Signatory

(Signature)
Naidu Babu Vanam
(Technical Manager/Quality Manager)

GLOBAL BIOMEDICAL SERVICES INDIA

10-35-5/1, 3rd floor, Near Federal Bank, Bhanoji Nagar, Visakhapatnam, Andhra Pradesh-530002.

+91 9247871397 | +91 9494320719 | 0891-2941397, Email id: globalbiomedicalservices@gmail.com, www.gbsindia.co.in

WE ARE PROVIDING CALIBRATION SERVICES ALL OVER INDIA



GLOBAL

BIO MEDICAL SERVICES INDIA



CALIBRATION CERTIFICATE

ULR - CC40412400000586F

Certificate Number : ESIC/P/5	Page no : 1 of 1
Service request no : ATPL/2024/10/0053	Date of Receipt : 15.08.2024
Service request date : 14.08.2024	Calibration done at : LAB
Calibration date : 15.08.2024	Suggested Date Of Next Calibration (As Requested By Customer) : 15.02.2025
Date of issue : 16.08.2024	

DETAILS OF CUSTOMER :-	M/s. Chief Medical Officer
	ESIC Model Hospital
	Bapunagar, Ahmedabad
	Gujarat-380024

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Nomenclature : Fixed Micropipette	Range : Fixed 1000 µl
Make : LABQUEST	Least Count : ---
Model : SF	Visual Inspection : Ok
Serial No : 696133	Accuracy/Acceptance Criteria : ---
ID Number :	Type : ---
Location : BIOCHEMISTRY	

STANDARD INSTRUMENTS USED FOR CALIBRATION

S.No	Nomenclature	Serial No	Identification No	Certificate No	Calibration Valid Upto
01	Electronic Semi Micro Balance	15114327	GBSI/MEC/WM-01	TVCSPL 24/02/0229-01	15.02.2025

Environmental Conditions	Calibration Procedure : SOP/MEC/004
Temperature : 25 ± 5°C	Reference Std : ISO 8655-6
Humidity : 35 to 85 % RH	

CALIBRATION RESULTS

Discipline :-Mechanical (Volume)

UUC Set Value	Measured Volume @ 27°C	Error	Uncertainty (±)	PASS/FAIL
µl	µl	µl	µl	
1000	1003.92	-3.92	0.93	PASS

* UUC Means Unit Under Calibration

* Reported Measured Volume at Average of 10 Measurements

Remarks :

- This certificate refers only to the particular item submitted for calibration.
- This certificate shall not be reproduced, except in full, without the written permission of Director of Global Biomedical Services .
- Results reported are valid at the time of & under the stated conditions of measurement.
- Laboratory Standards are Traceable to National and International Standards.
- The reported expanded uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by coverage factor k=2, which is for a normal distribution corresponds to a coverage probability and level of confidence of approximately 95%.

***** End Of Certificate *****

Calibrated by

Calibrated by

 V. Jyothi
 Calibration Engineer
 (Technical Manager)



Authorised Signatory

Authorised Signatory

 V. Naidu Babu Vanam
 (Technical Manager/Quality Manager)

GLOBAL BIOMEDICAL SERVICES INDIA

10-35-5/1, 3rd floor, Near Federal Bank, Bhanoji Nagar, Visakhapatnam, Andhra Pradesh-530002
 +91 9247871397 | +91 9494320719 | 0891-2941397, Email id: globalbiomedicalservices@gmail.com, www.gbsindia.co.in
 WE ARE PROVIDING CALIBRATION SERVICES ALL OVER INDIA