



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - OCTOBER 2024

LAB NO: 20400

LAB NAME: **GHSPL SAMBAJV KNJ HEALTH CARE LLP**

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 21/10/2024

Date of Result Published: 07/11/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1868	127	9.34	11.86	126.8 mg/dL	-0.02	0.55
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1374	31.84	14.13	4.5	32.3 mg/dL	0.1	0.24
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	873	1.04	15.81	0.17	1.2 mg/dL	0.94	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1712	5.76	16.85	0.97	5.9 mg/dL	0.14	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1531	5.8	9.91	0.58	5.66 g/dL	-0.24	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1492	3.45	10.24	0.35	3.55 g/dL	0.29	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1705	4.25	26.32	1.12	2.47 mg/dL	-1.59	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1858	124	10.76	13.35	120.3 mg/dL	-0.28	0.62
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1666	105	18	18.9	100 mg/dL	-0.26	0.93



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - OCTOBER 2024

LAB NO: 20400

LAB NAME: **GHSPL SAMBHAJV KNJ HEALTH CARE LLP**

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 21/10/2024

Date of Result Published: 07/11/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			

u* - method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

LAB ADDRESS

GHSPL SAMBHAJV KNJ HEALTH CARE LLP,
BHATJANGLA, KRISHNANAGAR, NH- 39, OPP
HOTEL HAVELI,
West Bengal

Coordinator Contact Details:

E-Mail: clinqc@cmcvellore.ac.in
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components.
***** End of Report *****

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.