

Urine Routine Analysis

ALL METHOD REPORT

Cycle - 13/2024

Date: 28/09/2024

Lab Code: 3700

Round No- 09

Parameters	Units	No. of Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	505	327	65%	1.020	1.020	Within Consensus
pH	-	505	332	66%	7.0	7.0	Within Consensus
Leukocytes	/µl	490	352	72%	2+3+	3+	Within Consensus
Nitrite	-	482	406	84%	Positive	.	-
Protein	mg/dl	504	385	76%	2+3+	3+	Within Consensus
Glucose	mg/dl	504	354	70%	3+4+	4+	Within Consensus
Ketones	-	500	321	64%	2+3+	3+	Within Consensus
Urobilinogen	mg/dl	496	354	71%	2+3+	3+	Within Consensus
Bilirubin	mg/dl	498	382	77%	3+	3+	Within Consensus
Blood (Hemoglobin)	/µl	500	373	75%	3+4+	3+	Within Consensus

Legends	≥ 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#)Late Result Submission
---------	-----------------	-----------------------	------------------	------------------	---------------------------

Chief Coordinator



Dr. Sanjay Mehrotra

Programme Director



Dr. Bandana Mehrotra

Prepared By: PS

Checked By: 

Doc. No.: ASS / FR / 10 / R 01 / Dt.: 05.01.2024