



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME

ALL ANALYSER REPORT - OCTOBER 2024

LAB NO: 2278

LAB NAME: PATHOCON DIAGNOSTICS

HbA1c

Analyte: HbA1c

S. No.	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
					CV	SDPA			
1	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	378	6	12.36	0.74	6.38 %	0.51	0.08

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

u* - method of uncertainty

LAB ADDRESS

PATHOCON DIAGNOSTICS,
At Nanjappas, No-10/A-1, 1st 'R' BLOCK, RAJAJI
NAGAR
Karnataka

Coordinator Contact Details:

E-Mail: clinqc@cmcvellore.ac.in
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components.
***** End of Report *****

Pamela Christudoss
Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.

Printed By: SUNIL
Lab no: 141665
PROFILE TEST (LFT) (00SE141665), LIVER FUNCTION TEST (LFT) (00
42 yrs / M
LIVER FUNCTION TEST (LFT) (00
LDL HDL
HDL DIR
6.8
LDL-cho
TOTAL BILL - 0.9
S ALBUMIN - 1.6
SGPT - 32
RBS - 83