

**Urine Routine Analysis**  
**METHOD WISE REPORT PART-02**

Cycle - 13/2024

Round No- 10

Date: 28/10/2024

Lab Code: 3115

Parameters	Units	Method Group	No. of Participants	Consensus		Consensus Result	Your Value	Remark
				Number	%			
Specific Gravity	-	Manual	151	139	92%	1.010-1.020	1.020	Within Consensus
pH	-	Manual	151	136	90%	6.0	6.0	Within Consensus
Leukocytes	/µl	Manual	142	134	94%	Negative	Negative	Within Consensus
Nitrite	-	Manual	140	136	97%	Negative	Negative	Within Consensus
Protein	mg/dl	Manual	151	144	95%	Negative	Negative	Within Consensus
Glucose	mg/dl	Manual	150	148	99%	Negative	Negative	Within Consensus
Ketones	-	Manual	149	147	99%	Negative	Negative	Within Consensus
Urobilinogen	mg/dl	Manual	146	140	96%	Negative	Negative	Within Consensus
Bilirubin	mg/dl	Manual	146	136	93%	Negative	Negative	Within Consensus
Blood (Hemoglobin)	/µl	Manual	147	142	97%	Negative	Negative	Within Consensus

Legends	≥ 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#)Late Result Submission
---------	-----------------	-----------------------	------------------	------------------	---------------------------

Chief Coordinator



Dr. Sanjay Mehrotra

Programme Director



Dr. Bandana Mehrotra

Prepared By: SSK

Checked By: 