

HORIBA

Explore the future

HORIBA India Private Limited

246, Okhla Industrial Estate, Phase-III,
New Delhi 110020, India

Tel : +91 (11) 4646 5000 / 4669 5001

Fax : +91 (11) 4646 5020 / 4669 5010

http : //www.horiba.com

Ref. No.: HIN/MED/2024-25/10010

Date: Jun 2, 2024

CALIBRATION CERTIFICATE

This is to certify that the Hematology analyzer Model: **Yumizen H550** bearing **S. No. 206YAXH03801** and installed at **LUPIN DIAGNOSTICS-VIJAYAWADA** was calibrated on **02/06/2024**.

The CV of repeatability check, Calibration and Control values were all found in range. Done the Calibration by using the **ABXMINOCAL** Calibrator Lot No. **CX494** Exp. Date – **05/06/2024**.

Calibration done and the instrument is working satisfactorily. Next Calibration Due On: June 1, 2025.



**For HORIBA MEDICAL,
HORIBA INDIA PVT. LIMITED.**