



Alere Medical Pvt. Ltd.  
No. 404, BPTP Park Centra,  
NH - 8, Sector - 30  
Opp. 32nd Milestone, Gurugram  
Haryana 122001, India

T: +91 124 4569000  
F: +91 124 4569011

### Calibration Certificate

**Instrument Name** : Urometer120  
**Serial Number** : 12210950  
  
**Customer Name** : Lupin Diagnostics ltd  
**Address** : Suryaraopeta  
Vijayawada-520002

**Standard used** : Self Calibration Mode  
**Standard Traceability** : In House Certificate

#### **Calibration**

The instrument function well in all the systems check, self check and maintenance Procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of  $333 \pm 10$ .

QC Performance : Acceptable  
Self Calibration : Acceptable

**Performance Qualification : Passed**

Calibration Date : 19/03/2024

**Calibration** : **Passed**

Calibration Valid up to : 18/03/2025

**For Alere Medical Pvt Ltd.**



U.Vinay satya kumar  
Engineer Technical Services  
M: 9949540344

REGISTERED OFFICE: 221 & 222, DLF TOWERS, SHIVAJI MARG, NAJAFGARH ROAD, NEW DELHI – 110015 (INDIA)  
TEL: +91-11-45089400  
CIN NO: - U74899DL1995PTC069456

[alere.com](http://alere.com)