



# Metropolis Healthcare Ltd., EQAS

MICROPATH LABORATORY (979)  
Outlier And Analyte Summary Report

Outlier Details For Cycle No 240207 and Sample No 01

Report Date : 18/10/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_


**Dated:** \_\_\_\_\_

Instrument : Clinitek status +

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	5.5	5,5,5,6
@ Specific Gravity	-	1.030	1.020,1.025,1.030
@ Glucose (urine)	-	Negative	Neg,Tr
@ Protein (Urine)	-	2+	+1,+2,+3
@ Ketone (urine)	-	Negative	Neg,Tr
@ Blood (urine)	-	Trace	Neg,Tr
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Absent	Abs

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Positive	Pos
Legend @ : Acceptable		Total Parameters	10
@ : Acceptable		Not Evaluated Parameters	0
X : Unacceptable		Evaluated Parameters	10
# : Not Evaluated		Outlier Parameters (X)	0
⌚ : Delayed Result Entry		EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kiroil Road, Kurla (W),  
Mumbai - 400070