

Calibration Certificate				
Product Name	Auto analyzer	Model	CSENSE 200 ^{CD}	
Tested Date	04/05/2024	SN	0532503025 IEC	
No.	Item	Standard and Demand	Obtained	Conclusion
1	Basic Function	Effective	Working Satisfactory	QC pass
2	Calibration and QC programming	Effective	Done	QC pass
3	Sample entry	Effective	Done	QC pass
4	All wavelength readings	20k to 50k	Working Satisfactory	QC pass
5	Reaction well temperature	37°C ± 0.3	37.1°C	QC pass
6	Reagent cooling	15°C lesser than RT	Checked	QC pass
7	Functionality check of mixer motor.	Judge whether the mixer is in the centre of the test cuvette/cleaning pool and whether you can hear the sound of rubbing up against the wall at the time of rotation	Checked	QC pass
8	Mechanical co-ordinate values	Instrument specific	Checked	QC pass
9	level sensor	Effective	Checked	QC pass
10	Collision sensor	Effective	Checked	QC pass
11	Input power source	210 to 240 V,50Hz	232 V, 50HZ	QC pass
Testing Result : Passed				
QC PASS				
Instrument calibrated as per the factory standard and the performed check values are satisfactory.				
Next calibration on: 03/05/2025				
Remarks:				
The Tester:		The Auditor:		



Medsource Ozone Biomedicals Pvt. Ltd.

Head Office: Plot No. 109, HSIIDC, Sector-31, Faridabad-121003, (Haryana), INDIA
 Registered Office: 238, Sant Nagar, East of Kailash, New Delhi-110065, INDIA
 Telephone: +91-129-4286600 | E-mail: mail@ozonebio.com | Website: www.ozonebio.com
 CIN: U85100DL2003PTC120213