



# Vaidyanatheshwara INSTRUMENTS PRIVATE LIMITED

## CERTIFICATE OF CALIBRATION



CC-2473



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer : M/s. PATHOCON DIAGNOSTICS Page 1 of 2

# 10/A-1, 1st 'R' Block, Rajajinagar  
Bengaluru - 560 010

### 2 Customer Reference

- 2.1 ULR No : CC247324100038800F
- 2.2 SRF No : 6121
- 2.3 Certificate No. : VI/24-25/6121-01
- 2.4 Format No. : VI-FRM-ME-073
- 2.5 Dc No & Dc Date : Refer Letter & 27-09-2024
- 2.6 Receipt Date : 28-09-2024
- 2.7 Date Of Issue : 30-09-2024

### 3 Details Of Device Under Calibration(DUC).

- 3.1 Nomenclature : Micro Pipette
- 3.2 Make & Model. : Thermo Scientific Finnpiptette F3
- 3.3 Range. : 20 - 200µl
- 3.4 SI No : QW03140
- 3.5 DUC Condition : Satisfactory
- 3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2022 E
- 3.7 No. of Pages : 2
- 3.8 Calibration Date : 30-09-2024
- 3.9 Calibration Due : 29-09-2025
- 3.10 Calibration done at : VI Volumetric Lab
- 3.11 Discipline : Mechanical (Mass & volume)

### 4 Environmental Condition

- 4.1 Temperature 22.4 - 22.8 °C Humidity 58.5 - 58.9 %RH

### 5 Standards Used for calibration

Sl. No.	Nomenclature	Make / Model	SL No / ID No	Range	Certificate. No.	Traceability to	Validity
1	Digital Weighing Balance	Radwag / XA 82/220.4Y	551565 / VI/ME/DWB-03	up to 220g	VI/24-25/INT-ME-118	VI Bangalore	10-09-2025
2	E2 Class Weights	LCGC / --	2020136 & VI-ME-E2C-001	1mg to 200g	VI/24-25-INT-ME-545-01	VI Bangalore	01-07-2025

### 6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. The Master's Used For Calibration are traceable to National standards/International Standards, Under Know Chain of Traceability
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By  
*Sanjay N Shrivastava*  
Sanjay N Shrivastava  
(Calibration Engineer)

Checked By  
*Ranjith Kumar P*  
Ranjith Kumar P  
(Lab. In Charge)





CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096. Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com



Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: VI/24-25/6121-01

Page 2 of 2

Range: 20 - 200µl

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty ±
	µl	µl	µl
20 - 200µl	20	20.21	0.05
	100	100.37	0.27
	200	200.59	1.3

Note:

1. Visual Inspection : Found Well.
2. Tripple Distilled water is used to Calibrate the MicroPipette.

Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

Sanjay N Shavi (Calibration Engineer)

Checked By

Ranjith Kumar P (Lab-In-Charge)



\*\*\*\*\*END OF CALIBRATION CERTIFICATE\*\*\*\*\*



# Vaidyanatheshwara INSTRUMENTS PRIVATE LIMITED

## CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Page 1 of 2

**1 Name and Address of the Customer** : M/s. PATHOCON DIAGNOSTICS  
# 10/A-1, 1st 'R' Block, Rajajinagar  
Bengaluru - 560 010

**2 Customer Reference**  
2.1 ULR No : CC247324100038801F  
2.2 SRF No : 6121  
2.3 Certificate No. : VI/24-25/6121-02  
2.4 Format No. : VI-FRM-ME-073  
2.5 Dc No & Dc Date : Refer Letter & 27-09-2024  
2.6 Receipt Date : 28-09-2024  
2.7 Date Of Issue : 30-09-2024

**3 Details Of Device Under Calibration(DUC).**  
3.1 Nomenclature : Micro Pipette  
3.2 Make : --  
3.3 Range. : 5 - 50µl  
3.4 SL No. : YE20CAS0103224  
3.5 DUC Condition : Satisfactory  
3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2022 E  
3.7 No.of Pages : 2  
3.8 Calibration Date : 30-09-2024  
3.9 Calibration Due : 29-09-2025  
3.10 Calibration done at : VI Volumetric Lab  
3.11 Discipline : Mechanical (Mass & volume)

**4 Environmental Condition**  
Temperature 21.6 - 21.9 °C Humidity 57.5 - 57.7 %RH

**5 Standards Used for calibration**

Sl. No.	Nomenclature	Make / Model	SL No / ID No	Range	Certificate. No.	Traceability to	Validity
1	Digital Weighing Balance	Radwag / XA 82/220.4Y	551565 / VI/ME/DWB-03	up to 220g	VI/24-25/INT-ME-118	VI Bangalore	10-09-2025
2	Digital Weighing Balance	Radwag / MYA5.4Y	544953 / VI/ME/DWB/02	up to 5g	VI/24-25/INT-ME-121	VI Bangalore	10-09-2025
3	E2 Class Weights	LCGC & --	2020136 & VI-ME-E2C-001	1mg to 200g	VI/24-25-INT-ME-545-01	VI Bangalore	01-07-2025

**6 Note:**

- The Calibration Certificate relates only to the above DUC
- Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- Corrections/erasing, invalidate the Calibration Certificate.
- The Master's Used For Calibration are traceable to National standards/International Standards, Under Know Chain of Traceability
- Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- Results Reported are valid at the time of and under the stated conditions of measurements.
- The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By  
  
Sanjay N Shavi  
(Calibration Engineer)

Checked By  
  
Ranji Kumar P  
(Lab- In-charge)



This report refers only to the item/gauge submitted and may not be reproduced except in full without written permission from Vaidyanatheshwara Instruments Private Limited Bangalore - 560096.



COMMITTED TO THE CUSTOMER SINCE - 1996

# Vaidyanatheshwara INSTRUMENTS PRIVATE LIMITED

## CERTIFICATE OF CALIBRATION



CC-2473



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: VI/24-25/6121-02

Page 2 of 2

Range: 5 - 50µl

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty ±
5 - 50µl	µl	µl	µl
	5	5.095	0.03
	20	20.42	0.05
	50	49.76	0.17

Note:

1. Visual Inspection : Found Well.
2. Triple Distilled water is used to Calibrate the MicroPipette.

Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

Sanjay N Shavi  
(Calibration Engineer)

Checked By

Ranjith Kumar P  
(Lab-In-Charge)



\*\*\*\*\*END OF CALIBRATION CERTIFICATE\*\*\*\*\*



CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096. Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com



Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer : M/s. PATHOCON DIAGNOSTICS # 10/A-1, 1st 'R' Block, Rajajinagar Bengaluru - 560 010

2 Customer Reference

- 2.1 ULR No : CC247324100038802F
2.2 SRF No : 6121
2.3 Certificate No. : VI/24-25/6121-03
2.4 Format No. : VI-FRM-ME-073
2.5 Dc No & Dc Date : Refer Letter & 27-09-2024
2.6 Receipt Date : 28-09-2024
2.7 Date Of Issue : 30-09-2024

3 Details Of Device Under Calibration(DUC).

- 3.1 Nomenclature : Micro Pipette
3.2 Make & Model. : Thermo Scientific Finnpiptette F3
3.3 Range. : 100 -1000µl
3.4 SL No : NW06666
3.5 DUC Condition : Satisfactory
3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2022 E
3.7 No.of Pages : 2
3.8 Calibration Date : 30-09-2024
3.9 Calibration Due : 29-09-2025
3.10 Calibration done at : VI Volumetric Lab
3.11 Discipline : Mechanical (Mass & volume)

4 Environmental Condition

Temperature 21.6 - 21.9 °C Humidity 57.5 - 57.7 %RH

5 Standards Used for calibration

Table with 8 columns: Sl. No., Nomenclature, Make / Model, SL No / ID No, Range, Certificate. No., Traceability to, Validity. Contains 2 rows of calibration data.

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
6.3. Corrections/erasing, invalidate the Calibration Certificate.
6.4. The Master's Used For Calibration are traceable to National standards/International Standards, Under Know Chain of Traceability
6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
6.6. Results Reported are valid at the time of and under the stated conditions of measurements
6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By Sanjay N Shaw (Calibration Engineer)

Checked By Ranjith Kumar P (Lab. In Charge)





COMMITTED TO THE  
CUSTOMER SINCE - 1996

# Vaidyanatheshwara Instruments Private Limited

## CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@vipigroup.com Web : www.vipigroup.com

Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: VI/24-25/6121-03

Page 2 of 2

Range: 100 -1000µl

Results:

Range	Cal Point µl	Volume observed µl	Expanded Uncertainty ± µl
100 -1000µl	100	100.34	0.27
	500	500.61	3.0
	1000	1000.78	5.9

**Note:**

1. Visual Inspection : Found Well.
2. Tripplle Distilled water is used to Calibrate the MicroPipette.

**Conclusion /Remarks:**

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

Sanjay N Shavi  
(Calibration Engineer)

Checked By

  
Ranjith Kumar P  
(Lab-In-Charge)

\*\*\*\*\*END OF CALIBRATION CERTIFICATE\*\*\*\*\*