

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Glucose	Mispa-Viva	115.20	mg/dl	4.21
# Bilirubin Direct	Mispa-Viva	0.28	mg/dl	0
! Albumin	Mispa-Viva	2.97	g/dL	-2.44
! Triglycerides	Mispa-Viva	107.60	mg/dl	-2.82
# HDL	Mispa-Viva	37.44	mg/dl	0
* ALT	Mispa-Viva	12.44	U/L	*
X ALP	Mispa-Viva	239.70	U/L	10.04
! CK	Mispa-Viva	20.27	U/L	-2.7
# LDH	Mispa-Viva	113.08	U/L	0
! Amylase	Mispa-Viva	78.85	U/L	2.16

Legend @ : z score ≤ 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score ≥ 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Mispa-Viva

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
X Glucose	mg/dl	115.20	85.00	4.21	2.98
@ Urea	mg/dl	23.11	23.56	-0.17	-0.31
@ Creatinine	mg/dl	3.83	3.42	-0.16	-1.12
# Bilirubin Direct	mg/dl	0.28	--		1.18
@ Total Protein	g/dL	7.02	--	0.46	0.67
! Albumin	g/dL	2.97	--	-2.44	-2.29
@ Uric Acid	mg/dl	4.65	5.18	-0.06	0.17
@ Calcium	mg/dl	8.84	--	0.30	0.25
@ Cholesterol	mg/dl	165.00	--	1.47	1.42
! Triglycerides	mg/dl	107.60	--	-2.82	-3.13
# HDL	mg/dl	37.44	--		0.32
* ALT	U/L	12.44	--	*	1.98
@ AST	U/L	19.69	--	0.74	0.72
X ALP	U/L	239.70	--	10.04	11.16
! CK	U/L	20.27	--	-2.70	2.51
# LDH	U/L	113.08	--		12.13
! Amylase	U/L	78.85	78.85	2.16	1.28
@ Lipase	U/L	63.67	65.84	0.44	-1.27

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	18
Not Evaluated Parameters	4
Evaluated Parameters	14
Outlier Parameters (X)	2
EQAS Score Biochemistry	85.71 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070