



Metropolis Healthcare Ltd., EQAS

AYUCARE PATHOLOGY LAB (1307)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240207 and Sample No 01

Report Date : 18/10/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual


Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	5	5,5,5,6
@ Specific Gravity	-	1.025	1.020,1.025,1.030
@ Glucose (urine)	-	Negative	Neg,Tr
@ Protein (Urine)	-	1+	+1,+2,+3
@ Ketone (urine)	-	Negative	Neg,Tr
@ Blood (urine)	-	Negative	Neg,Tr
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Absent	Abs
@ Leucocyte (Urine)	-	Negative	Neg,Tr

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Positive	Pos

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	11
Not Evaluated Parameters	0
Evaluated Parameters	11
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroil Road, Kurla (W),
 Mumbai - 400070