



Indian Calibration Services



An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechcs.com

CC-2146

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF REFRIGERATOR

Certificate No.:	24100018304	Issued On:	27/07/2024
ULR No.:	CC214624100018304F	Calibration Date:	24/07/2024
Job Identification No.:	ICS/C/T/07/18304	Next Calibration Date:	23/07/2025
Ref. No.:	SRF, Dated -22/07/2024		

CALIBRATED FOR:	M/s Tata 1mg Technologies Pvt. Ltd. Plot No. 16, 17 & 18, 3 rd Floor, JR Plaza District Shopping Centre Road, Lal Kothi, Jaipur, Rajasthan-302015.
------------------------	---

EQUIPMENT DESCRIPTION			
Name	Refrigerator	Visual Inspection	Ok
Make/Model	Celfrost / FKG400	I.D. No.	Tata1mg/JPR/LAB/REF-01
Serial No.	20138210300522	Least Count	
Range	2°C to 8°C	Calibration Site	On-Site
Location	Lab		

ENVIRONMENTAL CONDITIONS			
Temperature	25.0 °C ± 5.0 °C	Humidity	50 ± 15 % RH

STANDARD REFERENCE DETAILS						
Traceable to National / International Standard						
Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to	
Dig. Multichannel Data Logger	Countron	2410000018	Indian Calibration Services, Delhi	05/01/2024	05/01/2025	
USED EQUIPMENT DETAILS						
N.A.						
PRINCIPLE/METHODOLOGY OF CALIBRATION:						
As per Calibration Procedure No.: ICS/CAL/SOP-TH04 (ITS-90) (Comparison Method)						

RESULTS: Thermal Calibration (Temperature)

On Working Range:	Corrected Value of Standard (Average)	Uncertainty at approx 95% C.L and coverage factor k=2	Location Diagram
Location - 1 (L-1)	2.6 °C	± 1.1°C	
Location - 2 (L-2)	3.5 °C	± 1.1°C	
Location - 3 (L-3)	4.3 °C	± 1.1°C	
Location - 4 (L-4)	5.0 °C	± 1.1°C	

^The reported expanded uncertainty of measurement is stated as the standard uncertainty in measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

***END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of Indian Calibration Services

Checked by

Authorised Signatory

RAJESH KAPOOR
Operation-Head