

TMSCC

TESTING MACHINE SERVICE AND CALIBRATION CENTRE
Precision is Our Destination.....



An ISO 9001 : 2015 Certified company
Calibration of Various Instruments & Testing Machines...

CC - 3125

Form No.: TMSCC/R/23

CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE NO.: 2024/11/1287 Page 1 of 1

ISSUE DATE : 9-Nov-2024

ULR No.: CC312524000024143P

1.0 ISSUED TO : M/s.: Succes Diagnostic Centre
8C, Mohan Lal Street, Shyambazar,
(Near R.G. Kar Medical College), Kolkata - 700004.

1.1 Service Request Form No.: SRF/2024/11/07/05

1.2 Service Request Date : 7-Nov-2024

1.3 Location : At Lab

1.4 Description identification of item to be calibrated :

i Name :	Digital Thermometer	ii Make :	APTECHDEALS
iii Model / Type No.:	--	iv S/L No.:	--
v I.D.No.:	SDC/DTM/02	vi Job Code No.:	2024/11/1287
vii Range :	(-50 to 110)°C	viii Resolution :	0.1 °C
ix Accuracy :	As Per DKD R5 - 1	x End User :	--

1.5 Full / Partial Calibration : Partial Calibration

1.6 Applicable specification of item to be calibrated: Accuracy / permissible limit : Not Specified.

1.7 Date of receipt of item : 7-Nov-2024 **1.8** Date of calibration : 7-Nov-2024

1.9 Calibration due on : 6-Nov-2025 **2.0** Frequency of calibration once in : 12 Months

2.1 Environmental condition during calibration :
Temperature : 25.5 °C
Humidity : 56 % RH

2.2 Basis of calibration : SOP/10/02

2.3 Traceability : Standards used for calibration are traceable to National Standards through NABL accredited Laboratory.

Name of Instrument	Sl. No. / Id no.	Certificate No.	Lab Certificate No.	Calibrated On	Due On
Digital Temperature Indicator with Sensor(RTD)	(Indicator Sl. No.- 18K588073 (Sensor ID No.- TMSCC/RTD/01))	TSC/23-24/13010-4	CC-2231	09-11-2023	08-11-2024

CALIBRATION RESULTS

Sl. No.	Ref. Bath set in °C	Reading at DUC in °C	Observed Reading at Ref. Std. in °C (Avg. of five readings)	Error in °C	Uncertainty in ± °C	Acceptance Criteria
1	-35.0	-35.0	-34.940	0.060	0.18	Pass
2	-20.0	-20.0	-19.966	0.034	0.18	Pass
3	2.0	2.0	1.997	-0.003	0.18	Pass
4	10.0	10.0	9.983	-0.017	0.18	Pass
5	50.0	50.0	49.915	-0.085	0.31	Pass
6	110.0	110.0	109.813	-0.187	0.31	Pass

Maximum Permissible Error : ± 0.3% of rdg

Measurement Uncertainty at 95% Confidence Level where Coverage Factor k = 2

REMARKS : The DUC has been calibrated over its range. The readings observed are tabulated above. The reference standard is traceable to National standard.

DUC: Device Under Calibration.

Physical Status of the DUC : OK

Calibrated By :


R. Ghosh
(Calibration Engineer)

Checked By :

C. Ghosh
(Technical Director)

Approved By :

P. K. Modak
(Sr. Calibration Engineer)



END OF CERTIFICATE