

Urine Routine Analysis

METHOD WISE REPORT PART-01

Cycle - 13/2024

Lab Code: 3348

Round No- 10

Date: 28/10/2024

Parameters	Units	Method Group	No. of Participants	Consensus		Consensus Result	Your Value	Remark
				Number	%			
Specific Gravity	-	Semi-Automated	217	133	61%	1.020	1.020	Within Consensus
pH	-	Semi-Automated	217	206	95%	6.0	6.0	Within Consensus
Leukocytes	/µl	Semi-Automated	217	214	99%	Negative	Negative	Within Consensus
Nitrite	-	Semi-Automated	213	204	96%	Negative	Negative	Within Consensus
Protein	mg/dl	Semi-Automated	217	216	100%	Negative	Negative	Within Consensus
Glucose	mg/dl	Semi-Automated	217	215	99%	Negative	Negative	Within Consensus
Ketones	-	Semi-Automated	217	215	99%	Negative	Negative	Within Consensus
Urobilinogen	mg/dl	Semi-Automated	217	208	96%	Negative	Negative	Within Consensus
Bilirubin	mg/dl	Semi-Automated	217	215	99%	Negative	Negative	Within Consensus
Blood (Hemoglobin)	/µl	Semi-Automated	217	210	97%	Negative	Negative	Within Consensus

Legends	≥ 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#)Late Result Submission
---------	-----------------	-----------------------	------------------	------------------	---------------------------

Chief Coordinator



Dr. Sanjay Mehrotra


Programme Director



Dr. Bandana Mehrotra

Prepared By: SSK

Checked By:



Doc. No.: ASS / FR / 10 A / R 01 / Dt.: 05.01.2024