



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTIC LIMITED NALAGARH (1370)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240207 and Sample No 01

Report Date : 18/10/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
@ Sodium	ST 200 plus Electrolyte analyser	165.39	mmol/L	0
@ Potassium	ST 200 plus Electrolyte analyser	6.21	mmol/L	0


Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
⊙ : Delayed Result Entry  
\* : Not considered for evaluation.

Problem Classification: \_\_\_\_\_

Corrective Action: \_\_\_\_\_

Reviewed by: \_\_\_\_\_

Dated: \_\_\_\_\_

  
**KRSNAA DIAGNOSTICS LTD.**  
**CHC Nalagarh, (H.P.)**  
**PIN-174101**



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTIC LIMITED NALAGARH (1370)

## Outlier And Analyte Summary Report

Result Details For Cycle No 240207 and Sample No 01

Report Date : 18/10/2024

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	87.00	90.79	0.00	0.00
@ Urea	mg/dl	23.90	26.30	0.00	0.00
@ Creatinine	mg/dl	3.60	4.14	0.00	0.00
@ Bilirubin Total	mg/dl	0.22	0.46	0.00	0.00
@ Bilirubin Direct	mg/dl	0.03	0.13	0.00	0.00
@ Total Protein	g/dL	7.62	6.87	0.00	0.00
@ Albumin	g/dL	3.95	4.00	0.00	0.00
@ Uric Acid	mg/dl	4.60	4.56	0.00	0.00
@ Calcium	mg/dl	7.68	8.47	0.00	0.00
@ Cholesterol	mg/dl	145.00	156.62	0.00	0.00
@ Triglycerides	mg/dl	122.00	132.42	0.00	0.00
# HDL	mg/dl	10.00	7.62	0.00	0.00
# ALT	U/L	3.60	3.14	0.00	0.00
@ AST	U/L	13.40	15.72	0.00	0.00
@ ALP	U/L	83.00	78.17	0.00	0.00

Instrument : ST 200 plus Electrolyte analyser

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	106.15	109.26	0.00	0.00
@ Sodium	mmol/L	165.39	154.63	0.00	0.00
@ Potassium	mmol/L	6.21	5.22	0.00	0.00

Legend @ : Acceptable  
 : 2.0 < z score < 3.0 - Warning  
 X : z score ≥3 - Unacceptable  
 # : Not Evaluated  
 ⊙ : Delayed Result Entry  
 \* : Not considered for evaluation.

Total Parameters	18
Not Evaluated Parameters	2
Evaluated Parameters	16
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kirol Road, Kurla (W),  
 Mumbai - 400070

**KRSNAA DIAGNOSTICS LTD.**  
**CHC Nalagarh, (H.P.)**  
**PIN-174101**



**Metropolis Healthcare Ltd., EQAS**

KRSNAA DIAGNOSTIC LIMITED NALAGARH (1370)

**Outlier And Analyte Summary Report**

Outlier Details For Cycle No 240207 and Sample No 02

Report Date : 18/10/2024


Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Corrective Action: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Reviewed by: \_\_\_\_\_

Dated: \_\_\_\_\_

  
**KRSNAA DIAGNOSTICS LTD.**  
**GHC Nalagarh, (H.P.)**  
**PIN-174101**



**Metropolis Healthcare Ltd., EQAS**

KRSNAA DIAGNOSTIC LIMITED NALAGARH (1370)

Outlier And Analyte Summary Report

Result Details For Cycle No 240207 and Sample No 02

Report Date : 18/10/2024

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	87.00	79.72	0.00	0.00
@ Urea	mg/dl	28.10	29.05	0.00	0.00
@ Creatinine	mg/dl	1.13	1.20	0.00	0.00
@ Bilirubin Total	mg/dl	0.20	0.43	0.00	0.00
@ Bilirubin Direct	mg/dl	0.05	0.14	0.00	0.00
@ Total Protein	g/dL	7.84	7.41	0.00	0.00
@ Albumin	g/dL	4.28	4.09	0.00	0.00
@ Uric Acid	mg/dl	5.11	4.66	0.00	0.00
@ Calcium	mg/dl	8.83	8.77	0.00	0.00
@ Cholesterol	mg/dl	168.00	162.70	0.00	0.00
@ Triglycerides	mg/dl	131.00	130.79	0.00	0.00
@ HDL	mg/dl	36.80	32.82	0.00	0.00
@ ALT	U/L	9.30	5.21	0.00	0.00
@ AST	U/L	19.80	15.82	0.00	0.00
@ ALP	U/L	58.00	45.79	0.00	0.00

Instrument : ST 200 plus Electrolyte analyser

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	107.72	104.62	0.00	0.00
@ Sodium	mmol/L	142.74	136.97	0.00	0.00
@ Potassium	mmol/L	5.35	4.61	0.00	0.00

Legend @ : Acceptable  
 : 2.0 < z score < 3.0 - Warning  
 X : z score ≥ 3 - Unacceptable  
 # : Not Evaluated  
 ⊙ : Delayed Result Entry  
 \* : Not considered for evaluation.

Total Parameters	18
Not Evaluated Parameters	0
Evaluated Parameters	18
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

*Dr Puneet Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroh Road, Kurla (W),  
 Mumbai - 400070

*[Signature]*  
**KRSNAA DIAGNOSTICS LTD.**  
**CHC Nalagarh, (H.P.)**  
**PIN-174101**