



# A A CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 03985F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	240529.1 DT - 29 / 05 / 2024		ULR No.		CC2646240000 03985F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date		29.05.2024	
Address	CHC Pinjore-Swarghat Rd, Nalagarh-174101 (H.P.)			Recommended Date for Next Calib.		28.05.2025	
				Certificate Issue Dt.		01.06.2024	
				UUC Condition		Good	

### Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Mini Centrifuge M/c.	0 to 5000 RPM	100 RPM	Weswox 78437	KDHPNLCHCC06 Clinical Pathology

### Standard Equipments Used ( Traceable to National Standard )

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 12508 - 2018	RPM Measurement	55 %RH	24.6 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

### Calibration Result

Sl	Calib. Gauge Value	Master Gauge Value	Deviation	Uncertainty at approx 95% confidence	
	UUC ( in RPM )	( in RPM )	( in RPM )	Level and coverage factor $k = 2$ is $\pm$ (in RPM)	
1	100	100.6	-0.6	5.88	
2	500	501.3	-1.3	5.88	
3	1000	1002.5	-2.5	5.88	
4	2500	2503.9	-3.9	6.15	
5	3500	3505.2	-5.2	6.15	
6	5000	5007.6	-7.6	6.15	

### Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by  
  
Anand Chaurasia  
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012  
Ph. : 0120-4135529, Mob. : 9212172564, 9212172597  
E-mail : aacplinfo2005@gmail.com, accplinfo@rediffmail.com