



# AACALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 03986F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	240529.1	DT - 29 / 05 / 2024		ULR No.	CC2646240000 03986F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	29.05.2024		
Address	CHC Pinjore-Swarghat Rd, Nalagarh-174101 (H.P.)			Recommended Date for Next Calib.	28.05.2025		
				Certificate Issue Dt.	01.06.2024		
				UUC Condition	Good		

### Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Centrifuge M/c.	880 RPM	-----	Matrix BHCC-275/22-23	KDHPNLCHHM07 Hematology

### Standard Equipments Used ( Traceable to National Standard )

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 12508 - 2018	RPM Measurement	55 %RH	25.3 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08


### Calibration Result

SI	Calib. Gauge Value	Master Gauge Value	Deviation	Uncertainty at approx 95% confidence
	UUC ( in RPM )	( in RPM )	( in RPM )	Level and coverage factor $k = 2$ is $\pm$ (in RPM)
1	880	878.5	1.5	3.90

### Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----



Calibrated by

Approved by  
  
Anand Chaurasia  
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012  
Ph. : 0120-4135529, Mob. : 9212172564, 9212172597  
E-mail : aacplinfo2005@gmail.com, acplinfo@rediffmail.com