



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 04063F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	240531.1 DT - 31 / 05 / 2024			ULR No.	CC2646240000 04063F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	31.05.2024		
Address	CHC Pinjore-Swarghat Rd, Nalagarh-174101 (H.P.)			Recommended Date for Next Calib.	30.05.2025		
				Certificate Issue Dt.	01.06.2024		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	100 to 1000 µl	5 µl	Dragon Lab YE221AW0154709	KDHPNLCHBC15 Bio Chemsitry

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/744/JAN/24	04/03/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6 - 2022	Volume Measurement	54 %RH	23.8 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result


S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in µl)
1	100	100.03712	-0.03712	0.25
2	500	500.08526	-0.08526	0.25
3	1000	1000.14291	-0.14291	0.25

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by
Anand Chaurasia
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012
Ph. : 0120-4135529, Mob. : 9212172564, 9212172597
E-mail : aacplinfo2005@gmail.com, accplinfo@rediffmail.com