



CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Page No.: 1 of 2

1 Name and Address of the Customer

M/s.: Acharya Tulsi Diagnostic Centre & Dental Care
Jainam Building No. 58, M.K.K. Road,
Near Srirampuram Metro Station, Pillar No. 190, Bangalore - 21.

2 Customer Reference

2.1 ULR No. : CC247324500041396F
2.2 SRF No. : 0742
2.3 Certificate No. : VI/24-25/0742-03
2.4 Format No. : VI-FRM-ME-038
2.5 Date of Issue : 30-05-2024

3 Details Of Device Under Calibration (DUC).

3.1 Nomenclature : Microscope
3.2 Make : SUSWOX
3.3 Id.No. : MS-01
3.4 Calibration Procedure No. / Ref Doc : SOP-16-34 / ISO 10936-1
3.5 No.of Pages : 2
3.6 Calibration Date : 27-05-2024
3.7 Calibration Due : 26-05-2025
3.8 Calibration Done at : Onsite
3.9 Discipline : Mechanical
3.10 Instrument Location : Lab

4 Environmental Condition

Temperature 22.4 - 24.8 °C Humidity 50 - 55 %RH

5 Standards Used for calibration

Sl. No.	Nomenclature	Make & Model	Sl. No. / ID.No.	Cal Agency	Certificate No.	Validity
1	Steel Gauge Block Set	KCP & M122/1	10748 / VI/ME/SGB/002	VI, Bangalore	VI/23-24/INT-ME-209-01	03-11-2024

6 Note:

- The Calibration Certificate relates only to the above DUC.
- Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab.
- Corrections/erasing, invalidate the Calibration Certificate.
- Calibration of the DUC are traceable to National standards/International Standards
- Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- Results Reported are valid at the time of and under the stated conditions of measurements.
- The usage of NABL Symbol is as per NABL guidelines given in NABL-133

Calibrated By
Vidya Subramani
Vidya Subramani
(Calibration Engineer)

Checked By
Umesh D
Umesh D
(Onsite In-Charge)



CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No. VI/24-25/0742-03

Page No.: 2 of 2

Results:

A). CALIBRATION OF MAGNIFICATION

SI.No.	NOMINAL VALUE	Standard Slip Gauge Size Used (mm)	Nominal Value (mm)	Observed value	Observed magnification Error In %
1	10 X	10	100	99.98	-0.02%
				99.97	
				99.98	
2	100 X	1	100	99.99	-0.01%
				99.99	
				99.98	

Note:

- Determination of Magnifications of Microscope by Comparison method using Gauge Block Set.

Conclusion:

- Uncertainty of calibration at 95.45 % Confidence level and Coverage Factor k = 2 :
- Uncertainty of Measurement : ± 0.6 % For Magnification
- The Reported Results are valid only for the conditions of the received instruments/gauges at the time of and under the stated conditions of the calibration stated conditions of the calibration.

Calibrated By

Vidya Subramani
Vidyasubramani
(Calibration Engineer)

Checked By

F. Ganesh
Umesh D
(Onsite In-Charge)

Authorised By

