

**DR BELANI S BLUE SHEILD MEDICAL CLINIC**

**EQUIPMENT NAME & NO:C1**

**MODEL: REMI C854/6**

**SR.NO: ZGLN-31649**

**DATE:10.09.2024**

DISPLAY	READINGS ON TACHOMETER					MEAN	DIFFERENCE OF	95%	SD
READING							MEASUREMENT	PERCENTILE	
1	1611	1608	1594	1599	1593	1601	-1600	1610.4	8.2
2	2292	2305	2287	2296	2302	2296.4	-2294.4	2304.4	7.3
3	2984	2988	2976	2972	2977	2979.4	-2976.4	2987.2	6.5
4	3190	3215	3205	3198	3196	3200.8	-3196.8	3213	9.6
5	3864	3844	3859	3853	3866	3857.2	-3852.2	3865.6	8.9

**CALIBRATED ON:10.09.2024**

**CALIBRATION DUE ON: 09.03.2025**

**PERFORMED RUPALI GHULGHULE  
BY**

