

DR BELANI S BLUE SHEILD MEDICAL CLINIC

EQUIPMENT NAME & NO:C2

MODEL: REMI C854/8

SR.NO: HFLC-4302

DATE:10.09.2024

DISPLAY READING	READINGS ON TACHOMETER					MEAN	DIFFERENCE OF MEASUREMENT	95%	SD
								PERCENTILE	
1	1059	1062	1055	1070	1066	1062	-1061.4	1069.2	5.9
2	2296	2279	2272	2291	2283	2284	-2282.2	2295	9.5
3	2794	2792	2785	2796	2779	2789	-2786.2	2795.6	7.0
4	3292	3295	3293	3299	3296	3295	-3291	3298.4	2.7
5	4011	4002	4019	4001	4009	4008	-4003.4	4017.4	7.3

CALIBRATED ON:10.09.2024

CALIBRATION DUE ON: 09.03.2025

**PERFORMED RUPALI GHULGHULE
BY**

