

CERTIFICATE OF CALIBRATION/ VERIFICATION

Customer Name: DR. BELANI'S BLUE SHIELD MEDICAL CLINIC LLP
Model: Fully Hematology Analyzer-H360
Serial No: K10012134077
Calibration Date: 28-10-2024
Next Calibration due on before: 27-10-2025

This is to certify that the above mentioned instrument has been verified of calibration according to the procedures provided by the manufacturer -

General Parameters:

Parameter	Measured Value	Acceptable Range	
Line Voltage	230V	230 V +/- 5%	Pass
HGB gain value	4.57 v	4.2/4.8	Pass
Current source voltage	57.07	50/75	Pass
HGB blank voltage	4.56	4.2/4.8	Pass
Ambient temperatures	25.7	15/30	Pass

The result obtained is as per specification & tolerance range, the above calibration was done with meco Digital multimeter Model No - 108B+TRMS. The calibration certificate is enclosed herewith.


Technical Service Department
Transasia Bio-medical Limited

TRANSASIA BIO-MEDICALS LTD.

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072 Tel.: (022) 4030 9000, Fax: (022) 2857 3030
Email: transasia@transasia.co.in Website: www.transasia.co.in