

CERTIFICATE OF CALIBRATION/ VERIFICATION

Customer Name: DR. BELANI'S BLUE SHIELD MEDICAL CLINIC LLP
Model: Fully Automated Bio-Chemistry Analyzer-EM 200
Serial No: B151214
Calibration Date: 23-07-2024
Next Calibration due on before: 22-07-2025

This is to certify that the above mentioned instrument has been verified of calibration according to the procedures provided by the manufacturer -

General Parameters:

Parameter	Measured Value	Acceptable Range	
Line Voltage	230V	230 V +/- 5%	Pass
Logic Voltage	5.0 V	4.9 V to 5.1 V	Pass
Lamp Voltage	12.0 V	11.8 V to 12.2 V	Pass
Flow cell Temperature	37.0 deg	37 +/- 0.5	Pass
45 Cuvette blank value for for all filter	Enclose value sheet for all cuvette	Abs 0.02-0.2	Pass

TRANSASIA BIO-MEDICALS LTD.

CERTIFICATE OF CALIBRATION/ VERIFICATION

Customer Name: DR. BELANI'S BLUE SHIELD MEDICAL CLINIC LLP
Model: Fully Automated Bio-Chemistry Analyzer-EM 200
Serial No: B151214
Calibration Date: 23-07-2024
Next Calibration due on before: 22-07-2025

This is to certify that the above mentioned instrument has been verified of calibration according to the procedures provided by the manufacturer -

General Parameters:

Parameter	Measured Value	Acceptable Range	
Line Voltage	230V	230 V +/- 5%	Pass
Logic Voltage	5.0 V	4.9 V to 5.1 V	Pass
Lamp Voltage	12.0 V	11.8 V to 12.2 V	Pass
Flow cell Temperature	37.0 deg	37 +/- 0.5	Pass
45 Cuvette blank value for for all filter	Enclose value sheet for all cuvette	Abs 0.02-0.2	Pass

TRANSASIA BIO-MEDICALS LTD.