



BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	
Address:	
Province-Postal code	
Contact Name	

Calibration Certificate

Date <small>(dd mmmm, yyyy)</small>	Instrument Model	Instrument Serial Number
--	------------------	--------------------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Specified Range @ 1.5ml/Minute	Measured Value	
1.425 – 1.575 ml/minute		

Column Holder temperature Reference: Service Manual: Section 7.9

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Specified Range @ 35°C	Measured Value	
33.4°C – 36.4°C		

Detector log value Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	

The next calibration is due on	<small>(dd mmmm, yyyy)</small>
--------------------------------	--------------------------------

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory: