

Urine Routine Analysis

ALL METHOD REPORT

Cycle - 13/2024

Date: 28/10/2024

Lab Code: 3460

Round No -10

Parameters	Units	No. of Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	498	450	90%	1.010-1.020	*1.030	Outside Consensus
pH	-	498	457	92%	6.0	6.0	Within Consensus
Leukocytes	/µl	483	469	97%	Negative	Negative	Within Consensus
Nitrite	-	474	456	96%	Negative	Negative	Within Consensus
Protein	mg/dl	498	487	98%	Negative	Negative	Within Consensus
Glucose	mg/dl	496	488	98%	Negative	Negative	Within Consensus
Ketones	-	494	488	99%	Negative	Negative	Within Consensus
Urobilinogen	mg/dl	491	471	96%	Negative	Negative	Within Consensus
Bilirubin	mg/dl	491	473	96%	Negative	Negative	Within Consensus
Blood (Hemoglobin)	/µl	493	474	96%	Negative	Negative	Within Consensus

Legends	≥ 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#) Late Result Submission
---------	-----------------	-----------------------	------------------	------------------	----------------------------

Chief Coordinator

Programme Director



Dr. Sanjay Mehrotra



Dr. Bandana Mehrotra

Prepared By: R

Checked By: 

Doc. No.: ASS / FR / 10 / R 01 / Dt.: 05.01.2024